

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062718\
 Data File : VX002968.D
 Acq On : 28 Jun 2018 06:11
 Operator : JC/MD
 Sample : J3718-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-30

Quant Time: Jun 28 07:47:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	272064	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	382627	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	356075	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	204641	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	183928	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
35) Dibromofluoromethane	5.51	113	149380	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
50) Toluene-d8	8.72	98	534912	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
62) 4-Bromofluorobenzene	11.14	95	195170	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.52	142	2608	6.97	ug/l	# 95
11) Tert butyl alcohol	3.07	59	4692	6.26	ug/l	96
16) Acetone	2.45	43	19146	12.29	ug/l	99
19) Methyl tert-butyl Ether	3.21	73	31229	3.24	ug/l	96
22) Diisopropyl ether	3.87	45	6036	0.60	ug/l	93
25) 2-Butanone	4.71	43	3707	1.60	ug/l	91
65) Chlorobenzene	10.14	112	7410	0.82	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062718\
Data File : VX002968.D
Acq On : 28 Jun 2018 06:11
Operator : JC/MD
Sample : J3718-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-30

Quant Time: Jun 28 07:47:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

