

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062718\
 Data File : VX002971.D
 Acq On : 28 Jun 2018 07:31
 Operator : JC/MD
 Sample : J3691-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-11

Quant Time: Jun 28 08:00:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	271449	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	378803	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	350342	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	201319	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	180451	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
35) Dibromofluoromethane	5.51	113	147394	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
50) Toluene-d8	8.72	98	528999	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
62) 4-Bromofluorobenzene	11.14	95	188571	44.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.28%	
Target Compounds						
10) Methyl Iodide	2.51	142	2391	6.91	ug/l	96
16) Acetone	2.45	43	1905	1.23	ug/l	97
19) Methyl tert-butyl Ether	3.21	73	11908	1.24	ug/l	95
27) cis-1,2-Dichloroethene	4.60	96	1114	0.34	ug/l #	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062718\
Data File : VX002971.D
Acq On : 28 Jun 2018 07:31
Operator : JC/MD
Sample : J3691-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-11

Quant Time: Jun 28 08:00:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

