

Data File : VX002972.D
Acq On : 28 Jun 2018 07:57
Operator : JC/MD
Sample : J3691-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-17

Quant Time: Jul 04 03:39:16 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	266270	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	373762	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	352185	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	205032	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	179854	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
35) Dibromofluoromethane	5.51	113	147273	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
50) Toluene-d8	8.72	98	525232	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
62) 4-Bromofluorobenzene	11.14	95	190632	45.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.46%	
Target Compounds						
65) Chlorobenzene	10.14	112	5921	0.662	ug/l	96
83) tert-Butylbenzene	11.77	119	7184	0.627	ug/l	83
85) sec-Butylbenzene	11.95	105	27012	1.945	ug/l #	56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX002972.D
Acq On : 28 Jun 2018 07:57
Operator : JC/MD
Sample : J3691-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-17

Quant Time: Jul 04 03:39:16 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

