

Data File : VX002973.D
Acq On : 28 Jun 2018 08:23
Operator : JC/MD
Sample : J3691-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-19

Quant Time: Jul 04 03:38:21 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	277062	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	390485	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	363271	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	216013	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	188522	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
35) Dibromofluoromethane	5.51	113	153981	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
50) Toluene-d8	8.72	98	549014	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
62) 4-Bromofluorobenzene	11.14	95	200023	45.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.86%	
Target Compounds						
16) Acetone	2.45	43	4920	3.102	ug/l	92
19) Methyl tert-butyl Ether	3.21	73	5571	0.567	ug/l #	90
65) Chlorobenzene	10.14	112	4677	0.507	ug/l	97
85) sec-Butylbenzene	11.95	105	7766	0.531	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX002973.D
Acq On : 28 Jun 2018 08:23
Operator : JC/MD
Sample : J3691-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-19

Quant Time: Jul 04 03:38:21 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

