

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062718\
 Data File : VX002929.D
 Acq On : 27 Jun 2018 11:38
 Operator : JC/MD
 Sample : VX0627MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0627MBL01

Quant Time: Jun 28 03:43:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	248569	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	346691	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	314412	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	166553	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	168686	49.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.20%	
35) Dibromofluoromethane	5.51	113	132465	48.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.50%	
50) Toluene-d8	8.72	98	485965	48.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.94%	
62) 4-Bromofluorobenzene	11.14	95	166089	42.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.92%	


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062718\
Data File : VX002929.D
Acq On : 27 Jun 2018 11:38
Operator : JC/MD
Sample : VX0627MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0627MBL01

Quant Time: Jun 28 03:43:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 833

Delta R.T. -0.00 min

Lab File: VX002929.D

Acq: 27 Jun 2018 11:38

Instrument :

MSVOA_X

ClientSampleId :


VX0627MBL01

Tot Ion:168 Resp: 248569

Ion Ratio Lower Upper

168 100

99 51.4 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 49.10 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.01 min

Lab File: VX002929.D


Acq: 27 Jun 2018 11:38

Tgt Ion: 65 Resp: 168686

Ion Ratio Lower Upper

65 100

67 51.3 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07


60000


40000

20000

0

Time--> 5.90 6.00 6.10 6.20 6.30





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. -0.00 min

Lab File: VX002929.D

Acq: 27 Jun 2018 11:38

Instrument :

MSVOA_X

ClientSampleId :

VX0627MBL01


Tot Ion:114 Resp: 346691

Ion Ratio Lower Upper

114 100

63 20.0 0.0 41.4




88 15.1 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 48.25 ug/l

RT: 5.51 min Scan# 807

Delta R.T. -0.00 min

Lab File: VX002929.D

Acq: 27 Jun 2018 11:38


Tgt Ion:113 Resp: 132465

Ion Ratio Lower Upper

113 100

111 101.2 81.4 122.0



192 24.2 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.47 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX002929.D

Acq: 27 Jun 2018 11:38

Instrument :

MSVOA_X

ClientSampleId :


VX0627MBL01

Tot Ion: 98 Resp: 485965

Ion Ratio Lower Upper




98 100

100 63.4 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 42.96 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. 0.00 min

Lab File: VX002929.D

Acq: 27 Jun 2018 11:38


Tgt Ion: 95 Resp: 166089

Ion Ratio Lower Upper

95 100

174 96.6 0.0 187.4



176 93.7 0.0 181.0




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V





#63

Chlorobenzene-d5

Concen: 50.00 ug/l

RT: 10.12 min Scan# 1563

Delta R.T. -0.00 min

Lab File: VX002929.D

Acq: 27 Jun 2018 11:38

Instrument :

MSVOA_X

ClientSampleId :

VX0627MBL01


Tot Ion:117 Resp: 314412

Ion Ratio Lower Upper

117 100

82 53.7 45.0 67.4




119 32.0 25.8 38.6



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.08 min Scan# 1885

Delta R.T. -0.01 min

Lab File: VX002929.D

Acq: 27 Jun 2018 11:38


Tgt Ion:152 Resp: 166553

Ion Ratio Lower Upper

152 100

115 54.6 40.1 120.2

150 156.1 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

