

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062719\
 Data File : VX010502.D
 Acq On : 27 Jun 2019 20:23
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Jun 28 06:52:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	258721	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	383457	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	347726	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	178222	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	123294	50.83	ug/l	0.00
Spiked Amount			Recovery	=	101.66%	
35) Dibromofluoromethane	5.49	113	119861	51.09	ug/l	0.00
Spiked Amount			Recovery	=	102.18%	
50) Toluene-d8	8.71	98	451338	50.28	ug/l	0.00
Spiked Amount			Recovery	=	100.56%	
62) 4-Bromofluorobenzene	11.13	95	160949	49.67	ug/l	0.00
Spiked Amount			Recovery	=	99.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.20	85	80388	44.692	ug/l	99
3) Chloromethane	1.32	50	90647	43.279	ug/l	98
4) Vinyl Chloride	1.41	62	102812	44.816	ug/l	100
5) Bromomethane	1.64	94	55140	49.336	ug/l	98
6) Chloroethane	1.71	64	65128	48.119	ug/l	98
7) Trichlorofluoromethane	1.92	101	183031	50.329	ug/l	99
8) Diethyl Ether	2.19	74	65925	50.550	ug/l	99
9) 1,1,2-Trichlorotrifluoroet	2.38	101	105706	47.964	ug/l	99
10) Methyl Iodide	2.51	142	155331	49.387	ug/l	98
11) Tert butyl alcohol	3.05	59	141303	222.614	ug/l	97
12) 1,1-Dichloroethene	2.37	96	104451	46.577	ug/l	98
13) Acrolein	2.29	56	74710	177.428	ug/l	99
14) Allyl chloride	2.73	41	155777	48.929	ug/l	99
15) Acrylonitrile	3.14	53	315863	251.339	ug/l	99
16) Acetone	2.45	43	265495	218.385	ug/l	100
17) Carbon Disulfide	2.57	76	249211	41.422	ug/l	100
18) Methyl Acetate	2.78	43	122956	50.745	ug/l	99
19) Methyl tert-butyl Ether	3.20	73	354986	50.318	ug/l	98
20) Methylene Chloride	2.85	84	120342	48.101	ug/l	98
21) trans-1,2-Dichloroethene	3.17	96	114413	47.437	ug/l	97
22) Diisopropyl ether	3.86	45	333483	50.099	ug/l	88
23) Vinyl Acetate	3.82	43	1472136	256.001	ug/l	100
24) 1,1-Dichloroethane	3.70	63	191672	49.815	ug/l	100
25) 2-Butanone	4.67	43	434397	243.590	ug/l	98
26) 2,2-Dichloropropane	4.59	77	146851	43.935	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	138730	49.776	ug/l	98
28) Bromochloromethane	5.01	49	88166	47.925	ug/l	100
29) Tetrahydrofuran	5.13	42	279413	252.955	ug/l	100
30) Chloroform	5.21	83	206611	49.907	ug/l	99
31) Cyclohexane	5.58	56	165456	47.221	ug/l	97
32) 1,1,1-Trichloroethane	5.49	97	181066	49.532	ug/l	99
36) 1,1-Dichloropropene	5.80	75	145821	47.855	ug/l	100
37) Ethyl Acetate	4.84	43	163813	52.256	ug/l	99
38) Carbon Tetrachloride	5.79	117	162572	49.269	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062719\
 Data File : VX010502.D
 Acq On : 27 Jun 2019 20:23
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleID :
 VSTDCCC050EC

Quant Time: Jun 28 06:52:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.46	83	186415	46.541	ug/l	99
40) Benzene	6.14	78	469958	49.504	ug/l	100
41) Methacrylonitrile	5.04	41	86405	50.235	ug/l	98
42) 1,2-Dichloroethane	6.20	62	151331	52.368	ug/l	98
43) Isopropyl Acetate	6.45	43	267496	51.576	ug/l	99
44) Trichloroethene	7.21	130	139617	49.150	ug/l	98
45) 1,2-Dichloropropane	7.51	63	119076	49.670	ug/l	94
46) Dibromomethane	7.66	93	87228	50.713	ug/l	99
47) Bromodichloromethane	7.90	83	160481	50.911	ug/l	100
48) Methyl methacrylate	7.77	41	124088	51.473	ug/l	99
49) 1,4-Dioxane	7.74	88	69664	1003.121	ug/l	98
51) 4-Methyl-2-Pentanone	8.64	43	851700	256.763	ug/l	99
52) Toluene	8.79	92	306701	49.059	ug/l	100
53) t-1,3-Dichloropropene	9.04	75	174186	49.288	ug/l	100
54) cis-1,3-Dichloropropene	8.43	75	191319	49.816	ug/l	98
55) 1,1,2-Trichloroethane	9.21	97	131642	51.057	ug/l	99
56) Ethyl methacrylate	9.18	69	197377	50.534	ug/l	98
57) 1,3-Dichloropropane	9.37	76	204380	51.065	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.31	63	437138	235.655	ug/l	99
59) 2-Hexanone	9.49	43	649479	248.781	ug/l	99
60) Dibromochloromethane	9.58	129	146914	50.890	ug/l	100
61) 1,2-Dibromoethane	9.67	107	142022	51.032	ug/l	100
64) Tetrachloroethene	9.34	164	136463	48.520	ug/l	99
65) Chlorobenzene	10.13	112	348903	49.423	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.22	131	131548	51.157	ug/l	98
67) Ethyl Benzene	10.25	91	596298	50.033	ug/l	100
68) m/p-Xylenes	10.36	106	458678	99.326	ug/l	100
69) o-Xylene	10.69	106	223652	49.248	ug/l	99
70) Styrene	10.71	104	386238	50.714	ug/l	99
71) Bromoform	10.85	173	112213	50.181	ug/l	98
73) Isopropylbenzene	11.02	105	598439	49.535	ug/l	99
74) N-amyl acetate	10.90	43	234164	51.287	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.27	83	199535	49.871	ug/l	99
76) 1,2,3-Trichloropropane	11.29	75	220026	67.749	ug/l	80
77) Bromobenzene	11.26	156	164203	48.675	ug/l	99
78) n-propylbenzene	11.36	91	655149	49.342	ug/l	100
79) 2-Chlorotoluene	11.42	91	386793	48.920	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	491777	49.377	ug/l	100
81) trans-1,4-Dichloro-2-buten	11.07	75	62420	44.838	ug/l	97
82) 4-Chlorotoluene	11.51	91	442178	49.148	ug/l	98
83) tert-Butylbenzene	11.77	119	494524	49.471	ug/l	100
84) 1,2,4-Trimethylbenzene	11.80	105	504060	50.375	ug/l	99
85) sec-Butylbenzene	11.94	105	586231	49.947	ug/l	99
86) p-Isopropyltoluene	12.06	119	537709	49.530	ug/l	100
87) 1,3-Dichlorobenzene	12.02	146	285793	48.534	ug/l	99
88) 1,4-Dichlorobenzene	12.10	146	285969	47.703	ug/l	100
89) n-Butylbenzene	12.38	91	453674	49.178	ug/l	100
90) Hexachloroethane	12.60	117	93174	49.635	ug/l	100
91) 1,2-Dichlorobenzene	12.39	146	282107	48.203	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	13.00	75	41355	47.940	ug/l	97


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062719\
Data File : VX010502.D
Acq On : 27 Jun 2019 20:23
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Quant Time: Jun 28 06:52:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.65	180	199577	49.174	ug/l	99
94) Hexachlorobutadiene	13.78	225	96396	48.597	ug/l	98
95) Naphthalene	13.83	128	629968	50.509	ug/l	100
96) 1,2,3-Trichlorobenzene	14.02	180	199722	49.494	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion:168 Resp: 258721

Ion Ratio Lower Upper

168 100

99 46.7 35.5 53.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

100000

80000

60000



40000

20000

0

Time-->

5.50 5.60 5.70 5.80



#2

Dichlorodifluoromethane

Concen: 44.692 ug/l

RT: 1.20 min Scan# 7

Delta R.T. 0.00 min

Lab File: VX010502.D


Acq: 27 Jun 2019 20:23

Tgt Ion: 85 Resp: 80388

Ion Ratio Lower Upper

85 100

87 32.9 16.3 48.8



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0

1.20

80000

60000


40000


20000

0

Time-->

1.15 1.20 1.25 1.30





#3

Chloromethane

Concen: 43.279 ug/l

RT: 1.32 min Scan# 28

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 50 Resp: 90647

Ion Ratio Lower Upper




50 100

52 34.3 26.5 39.7



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0



#4

Vinyl Chloride

Concen: 44.816 ug/l

RT: 1.41 min Scan# 42

Delta R.T. 0.00 min

Lab File: VX010502.D


Acq: 27 Jun 2019 20:23

Tgt Ion: 62 Resp: 102812

Ion Ratio Lower Upper



62 100


64 32.5 26.2 39.4



Abundance Ion 61.90 (61.60 to 62.60): VX0

Ion 63.90 (63.60 to 64.60): VX0





#5

Bromomethane

Concen: 49.336 ug/l

RT: 1.64 min Scan# 79

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 94 Resp: 55140

Ion Ratio Lower Upper




94 100

96 96.2 75.5 113.3



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0



#6

Chloroethane

Concen: 48.119 ug/l

RT: 1.71 min Scan# 92

Delta R.T. 0.00 min

Lab File: VX010502.D


Acq: 27 Jun 2019 20:23

Tgt Ion: 64 Resp: 65128

Ion Ratio Lower Upper



64 100


66 32.0 26.6 40.0



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0






#7




Trichlorofluoromethane
 Concen: 50.329 ug/l
 RT: 1.92 min Scan# 126
 Delta R.T. -0.01 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion:101 Resp: 183031
 Ion Ratio Lower Upper
 101 100
 103 64.3 52.3 78.5




Abundance Ion 100.90 (100.60 to 101.60): V
 Ion 102.80 (102.50 to 103.50): V





#8


Diethyl Ether
 Concen: 50.550 ug/l
 RT: 2.19 min Scan# 170
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Tgt Ion: 74 Resp: 65925
 Ion Ratio Lower Upper
 74 100
 45 88.8 44.8 134.4



Abundance Ion 74.00 (73.70 to 74.70): VX0
 Ion 45.00 (44.70 to 45.70): VX0









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 47.964 ug/l
 RT: 2.38 min Scan# 201
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion:101 Resp: 105706
 Ion Ratio Lower Upper
 101 100
 85 41.4 33.7 50.5
 151 88.5 70.1 105.1




Abundance Ion 100.90 (100.60 to 101.60): V
 Ion 84.90 (84.60 to 85.60): VX0
 Ion 150.80 (150.50 to 151.50): V






#10
 Methyl Iodide
 Concen: 49.387 ug/l
 RT: 2.51 min Scan# 222
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Tgt Ion:142 Resp: 155331
 Ion Ratio Lower Upper
 142 100
 127 40.2 31.1 46.7
 141 14.3 11.2 16.8



Abundance Ion 141.80 (141.50 to 142.50): V
 Ion 126.80 (126.50 to 127.50): V
 Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
Concen: 222.614 ug/l
RT: 3.05 min Scan# 311
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tot Ion: 59 Resp: 141303
Ion Ratio Lower Upper
59 100
57 10.7 7.7 11.5




#12

1,1-Dichloroethene
Concen: 46.577 ug/l
RT: 2.37 min Scan# 200
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion: 96 Resp: 104451
Ion Ratio Lower Upper
96 100
61 134.3 109.2 163.8
98 63.1 51.1 76.7





#13

Acrolein

Concen: 177.428 ug/l

RT: 2.29 min Scan# 187

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion: 56 Resp: 74710

Ion Ratio Lower Upper

56 100

55 69.9 56.9 85.3



#14

Allyl chloride

Concen: 48.929 ug/l

RT: 2.73 min Scan# 258

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23




Tgt Ion: 41 Resp: 155777


Ion Ratio Lower Upper

41 100

39 63.2 50.3 75.5

76 38.4 31.3 46.9





#15

Acrylonitrile

Concen: 251.339 ug/l

RT: 3.14 min Scan# 326

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :


MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tot Ion: 53 Resp: 315863

Ion	Ratio	Lower	Upper
53	100		
52	82.4	65.5	98.3
51	35.0	28.2	42.4






Abundance

Ion 53.00 (52.70 to 53.70): VX0

Ion 52.00 (51.70 to 52.70): VX0

Ion 51.00 (50.70 to 51.70): VX0



#16

Acetone

Concen: 218.385 ug/l

RT: 2.45 min Scan# 212


Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Tgt Ion: 43 Resp: 265495



Ion	Ratio	Lower	Upper
43	100		
58	36.0	28.9	43.3




Abundance

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0





#17

Carbon Disulfide

Concen: 41.422 ug/l

RT: 2.57 min Scan# 232

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 76 Resp: 249211

Ion Ratio Lower Upper

76 100

78 9.1 7.3 10.9



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0

2.57

150000



100000

50000

0

2.50 2.60 2.70

Time-->



#18

Methyl Acetate

Concen: 50.745 ug/l

RT: 2.78 min Scan# 266

Delta R.T. 0.00 min

Lab File: VX010502.D


Acq: 27 Jun 2019 20:23

Tgt Ion: 43 Resp: 122956

Ion Ratio Lower Upper

43 100

74 27.4 21.3 31.9



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0

80000

60000


40000


20000

0

2.70 2.75 2.80 2.85

Time-->









#19
Methvl tert-butvl Ether
Concen: 50.318 ug/l
RT: 3.20 min Scan# 335
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 73 Resp: 354986
Ion Ratio Lower Upper
73 100
57 21.5 16.4 24.6



#20
Methylene Chloride
Concen: 48.101 ug/l
RT: 2.85 min Scan# 279
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion: 84 Resp: 120342
Ion Ratio Lower Upper
84 100
49 106.4 87.5 131.3
51 34.4 27.7 41.5
86 65.1 53.4 80.0





#21

trans-1,2-Dichloroethene

Concen: 47.437 ug/l

RT: 3.17 min Scan# 330

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 96 Resp: 114413

Ion Ratio Lower Upper




96 100

61 115.7 95.6 143.4

98 64.3 52.5 78.7



Abundance Ion 95.90 (95.60 to 96.60): VX0
Ion 60.90 (60.60 to 61.60): VX0
Ion 97.90 (97.60 to 98.60): VX0



#22

Diisopropyl ether

Concen: 50.099 ug/l

RT: 3.86 min Scan# 444

Delta R.T. 0.01 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Tgt Ion: 45 Resp: 333483


Ion Ratio Lower Upper

45 100



43 64.4 63.3 94.9


87 34.1 25.8 38.8

59 12.2 9.6 14.4



Abundance Ion 45.00 (44.70 to 45.70): VX0
Ion 43.00 (42.70 to 43.70): VX0
Ion 87.00 (86.70 to 87.70): VX0
Ion 59.00 (58.70 to 59.70): VX0





#23

Vinyl Acetate

Concen: 256.001 ug/l

RT: 3.82 min Scan# 437

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 43 Resp: 1472136

Ion Ratio Lower Upper




43 100

86 12.3 9.9 14.9



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 86.00 (85.70 to 86.70): VX0



#24

1,1-Dichloroethane

Concen: 49.815 ug/l

RT: 3.70 min Scan# 417

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23


Tgt Ion: 63 Resp: 191672

Ion Ratio Lower Upper

63 100

98 7.7 3.9 11.7



100 4.6 2.4 7.1




Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 97.90 (97.60 to 98.60): VX0

Ion 99.90 (99.60 to 100.60): VX0





#25

2-Butanone

Concen: 243.590 ug/l

RT: 4.67 min Scan# 577

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 43 Resp: 434397

Ion Ratio Lower Upper




43 100

72 30.2 23.3 34.9



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 72.00 (71.70 to 72.70): VX0



#26

2,2-Dichloropropane

Concen: 43.935 ug/l

RT: 4.59 min Scan# 563

Delta R.T. 0.01 min

Lab File: VX010502.D


Acq: 27 Jun 2019 20:23

Tgt Ion: 77 Resp: 146851

Ion Ratio Lower Upper



77 100


97 25.8 13.1 39.3



Abundance Ion 76.90 (76.60 to 77.60): VX0

Ion 96.90 (96.60 to 97.60): VX0









#27

cis-1,2-Dichloroethene
Concen: 49.776 ug/l
RT: 4.60 min Scan# 565
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tot Ion	Ratio	Lower	Upper
96	100		
61	122.7	0.0	251.8
98	64.7	0.0	129.4




#28

Bromochloromethane
Concen: 47.925 ug/l
RT: 5.01 min Scan# 633
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion	Ratio	Lower	Upper
49	100		
129	3.0	0.0	5.6
130	101.3	81.4	122.0





#29

Tetrahydrofuran

Concen: 252.955 ug/l

RT: 5.13 min Scan# 652

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 42 Resp: 279413

Ion Ratio Lower Upper

42 100

72 50.9 40.6 60.8




71 47.3 37.8 56.8



Abundance Ion 42.00 (41.70 to 42.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

Ion 71.00 (70.70 to 71.70): VX0



#30

Chloroform

Concen: 49.907 ug/l

RT: 5.21 min Scan# 666

Delta R.T. 0.01 min

Lab File: VX010502.D


Acq: 27 Jun 2019 20:23

Tgt Ion: 83 Resp: 206611

Ion Ratio Lower Upper



83 100


85 64.1 50.9 76.3



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0





#31

Cyclohexane

Concen: 47.221 ug/l

RT: 5.58 min Scan# 726

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tot Ion: 56 Resp: 165456

Ion Ratio Lower Upper

56 100

69 30.8 25.9 38.9

84 94.8 78.0 117.0



#32

1,1,1-Trichloroethane

Concen: 49.532 ug/l

RT: 5.49 min Scan# 712

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23




Tgt Ion: 97 Resp: 181066


Ion Ratio Lower Upper

97 100

99 65.5 51.8 77.8

61 39.8 32.1 48.1









#33
 1,2-Dichloroethane-d4
 Concen: 50.827 ug/l
 RT: 6.06 min Scan# 805
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 65 Resp: 123294
 Ion Ratio Lower Upper
 65 100
 67 54.8 0.0 110.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Tgt Ion: 114 Resp: 383457
 Ion Ratio Lower Upper
 114 100
 63 17.4 0.0 33.8
 88 14.8 0.0 27.6









#35
 Dibromofluoromethane
 Concen: 51.089 ug/l
 RT: 5.49 min Scan# 712
 Delta R.T. -0.01 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion:113 Resp: 119861
 Ion Ratio Lower Upper
 113 100
 111 102.4 81.2 121.8
 192 22.9 18.6 27.8




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V






#36
 1,1-Dichloropropene
 Concen: 47.855 ug/l
 RT: 5.80 min Scan# 762
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Tgt Ion: 75 Resp: 145821
 Ion Ratio Lower Upper
 75 100
 110 41.2 20.8 62.2
 77 31.4 25.4 38.0



Abundance Ion 74.90 (74.60 to 75.60): VX0
 Ion 109.90 (109.60 to 110.60): V
 Ion 76.90 (76.60 to 77.60): VX0





#37

Ethyl Acetate

Concen: 52.256 ug/l

RT: 4.84 min Scan# 604

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 43 Resp: 163813

Ion Ratio Lower Upper

43 100

61 15.3 12.4 18.6




70 11.9 9.8 14.8



Abundance Ion 43.00 (42.70 to 43.70): VX010502.D (-555) (-)

Ion 60.90 (60.60 to 61.60): VX010502.D (-555) (-)

Ion 70.00 (69.70 to 70.70): VX010502.D (-555) (-)



#38

Carbon Tetrachloride

Concen: 49.269 ug/l

RT: 5.79 min Scan# 760

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23


Tgt Ion: 117 Resp: 162572

Ion Ratio Lower Upper

117 100

119 95.0 78.2 117.4



121 31.6 24.7 37.1




Abundance Ion 116.80 (116.50 to 117.50): VX010502.D (-711) (-)

Ion 118.80 (118.50 to 119.50): VX010502.D (-711) (-)

Ion 120.80 (120.50 to 121.50): VX010502.D (-711) (-)









#39
Methylvlcyclohexane
Concen: 46.541 ug/l
RT: 7.46 min Scan# 1035
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 83 Resp: 186415
Ion Ratio Lower Upper
83 100
55 67.1 53.0 79.6
98 46.6 37.8 56.8



#40
Benzene
Concen: 49.504 ug/l
RT: 6.14 min Scan# 818
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion: 78 Resp: 469958
Ion Ratio Lower Upper
78 100
77 23.5 18.6 28.0









#41
Methacrylonitrile
Concen: 50.235 ug/l
RT: 5.04 min Scan# 638
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion:	41	Resp:	86405
Ion	Ratio	Lower	Upper
41	100		
39	54.6	43.8	65.8
67	84.9	66.6	100.0
52	33.6	24.5	36.7




Abundance Ion 41.00 (40.70 to 41.70): VX0
Ion 39.00 (38.70 to 39.70): VX0
Ion 67.00 (66.70 to 67.70): VX0
Ion 52.00 (51.70 to 52.70): VX0






#42
1,2-Dichloroethane
Concen: 52.368 ug/l
RT: 6.20 min Scan# 827
Delta R.T. 0.01 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion:	62	Resp:	151331
Ion	Ratio	Lower	Upper
62	100		
98	10.5	0.0	22.2



Abundance Ion 61.90 (61.60 to 62.60): VX0
Ion 97.90 (97.60 to 98.60): VX0





#43

Isopropyl Acetate

Concen: 51.576 ug/l

RT: 6.45 min Scan# 868

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 43 Resp: 267496

Ion Ratio Lower Upper

43 100

61 21.4 17.8 26.6




87 15.6 13.0 19.4



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#44

Trichloroethene

Concen: 49.150 ug/l

RT: 7.21 min Scan# 994

Delta R.T. 0.00 min

Lab File: VX010502.D


Acq: 27 Jun 2019 20:23

Tgt Ion:130 Resp: 139617

Ion Ratio Lower Upper



130 100


95 92.2 0.0 180.8



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VX0









#45
 1,2-Dichloropropane
 Concen: 49.670 ug/l
 RT: 7.51 min Scan# 1043
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 63 Resp: 119076
 Ion Ratio Lower Upper
 63 100
 65 34.1 24.8 37.2



#46
 Dibromomethane
 Concen: 50.713 ug/l
 RT: 7.66 min Scan# 1067
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Tgt Ion: 93 Resp: 87228
 Ion Ratio Lower Upper
 93 100
 95 83.5 67.5 101.3
 174 118.0 95.8 143.6





#47

Bromodichloromethane

Concen: 50.911 ug/l

RT: 7.90 min Scan# 1106

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 83 Resp: 160481

Ion Ratio Lower Upper




83 100

85 64.5 51.5 77.3

127 9.7 7.8 11.8



Abundance Ion 82.90 (82.60 to 83.60): VX0
Ion 84.90 (84.60 to 85.60): VX0
Ion 126.80 (126.50 to 127.50): V



#48

Methyl methacrylate

Concen: 51.473 ug/l

RT: 7.77 min Scan# 1085

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23


Tgt Ion: 41 Resp: 124088

Ion Ratio Lower Upper



41 100


69 96.1 76.1 114.1

39 60.9 47.6 71.4



Abundance Ion 41.00 (40.70 to 41.70): VX0
Ion 69.00 (68.70 to 69.70): VX0
Ion 39.00 (38.70 to 39.70): VX0









#49
1,4-Dioxane
Concen: 1003.121 ug/l
RT: 7.74 min Scan# 1080
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion: 88 Resp: 69664
Ion Ratio Lower Upper
88 100
43 26.4 20.2 30.2
58 60.1 49.5 74.3




Abundance Ion 88.00 (87.70 to 88.70): VX0
Ion 43.00 (42.70 to 43.70): VX0
Ion 58.00 (57.70 to 58.70): VX0






#50
Toluene-d8
Concen: 50.284 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion: 98 Resp: 451338
Ion Ratio Lower Upper
98 100
100 64.3 51.6 77.4



Abundance Ion 98.00 (97.70 to 98.70): VX0
Ion 100.00 (99.70 to 100.70): V









#51
 4-Methyl-2-Pentanone
 Concen: 256.763 ug/l
 RT: 8.64 min Scan# 1228
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 43 Resp: 851700
 Ion Ratio Lower Upper
 43 100
 58 41.1 33.5 50.3



#52
 Toluene
 Concen: 49.059 ug/l
 RT: 8.79 min Scan# 1252
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Tgt Ion: 92 Resp: 306701
 Ion Ratio Lower Upper
 92 100
 91 174.9 139.8 209.6





#53

t-1,3-Dichloropropene

Concen: 49.288 ug/l

RT: 9.04 min Scan# 1294

Delta R.T. 0.01 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion: 75 Resp: 174186

Ion Ratio Lower Upper

75 100

77 31.6 25.4 38.2



#54

cis-1,3-Dichloropropene

Concen: 49.816 ug/l

RT: 8.43 min Scan# 1194

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23




Tgt Ion: 75 Resp: 191319


Ion Ratio Lower Upper

75 100

77 33.2 25.2 37.8

39 43.5 34.5 51.7









#55
 1,1,2-Trichloroethane
 Concen: 51.057 ug/l
 RT: 9.21 min Scan# 1322
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion:	97	Resp:	131642
Ion	Ratio	Lower	Upper
97	100		
83	84.2	67.4	101.0
85	55.9	43.8	65.8
99	63.5	50.3	75.5




Abundance Ion 96.90 (96.60 to 97.60): VX0
 Ion 82.90 (82.60 to 83.60): VX0
 Ion 84.90 (84.60 to 85.60): VX0
 Ion 98.90 (98.60 to 99.60): VX0






#56
 Ethyl methacrylate
 Concen: 50.534 ug/l
 RT: 9.18 min Scan# 1316
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Tgt Ion:	69	Resp:	197377
Ion	Ratio	Lower	Upper
69	100		
41	60.1	47.1	70.7
39	34.6	26.6	40.0



Abundance Ion 69.00 (68.70 to 69.70): VX0
 Ion 41.00 (40.70 to 41.70): VX0
 Ion 39.00 (38.70 to 39.70): VX0





#57

1,3-Dichloropropane

Concen: 51.065 ug/l

RT: 9.37 min Scan# 1348

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 76 Resp: 204380

Ion Ratio Lower Upper

76 100

78 32.8 26.1 39.1



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0

9.37

150000



100000

50000

0

Time-->

9.30 9.35 9.40 9.45



#58

2-Chloroethyl Vinyl ether

Concen: 235.655 ug/l

RT: 8.31 min Scan# 1174

Delta R.T. 0.00 min

Lab File: VX010502.D


Acq: 27 Jun 2019 20:23

Tgt Ion: 63 Resp: 437138

Ion Ratio Lower Upper

63 100

106 31.2 24.7 37.1



Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 105.90 (105.60 to 106.60): V

8.31

300000


200000


100000

0

Time-->

8.20 8.25 8.30 8.35 8.40









#59
2-Hexanone
Concen: 248.781 ug/l
RT: 9.49 min Scan# 1368
Delta R.T. 0.01 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 43 Resp: 649479
Ion Ratio Lower Upper
43 100
58 57.6 29.0 87.1



#60
Dibromochloromethane
Concen: 50.890 ug/l
RT: 9.58 min Scan# 1382
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion: 129 Resp: 146914
Ion Ratio Lower Upper
129 100
127 76.5 38.4 115.1









#61
 1,2-Dibromoethane
 Concen: 51.032 ug/l
 RT: 9.67 min Scan# 1397
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 107 Resp: 142022
 Ion Ratio Lower Upper
 107 100
 109 94.8 75.8 113.6



#62
 4-Bromofluorobenzene
 Concen: 49.668 ug/l
 RT: 11.13 min Scan# 1637
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Tgt Ion: 95 Resp: 160949
 Ion Ratio Lower Upper
 95 100
 174 93.0 0.0 187.6
 176 91.6 0.0 184.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion:117 Resp: 347726
Ion Ratio Lower Upper
117 100
82 51.0 41.2 61.8
119 31.9 25.5 38.3



#64
Tetrachloroethene
Concen: 48.520 ug/l
RT: 9.34 min Scan# 1342
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion:164 Resp: 136463
Ion Ratio Lower Upper
164 100
166 131.6 103.6 155.4
129 90.1 73.0 109.4
131 89.5 71.0 106.4









#65
Chlorobenzene
Concen: 49.423 ug/l
RT: 10.13 min Scan# 1473
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion:112 Resp: 348903
Ion Ratio Lower Upper
112 100
114 32.6 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 51.157 ug/l
RT: 10.22 min Scan# 1487
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion:131 Resp: 131548
Ion Ratio Lower Upper
131 100
133 96.8 47.7 143.1
119 66.4 32.6 97.7









#67
Ethyl Benzene
Concen: 50.033 ug/l
RT: 10.25 min Scan# 1492
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 91 Resp: 596298
Ion Ratio Lower Upper
91 100
106 32.8 26.2 39.2



#68
m/p-Xylenes
Concen: 99.326 ug/l
RT: 10.36 min Scan# 1510
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion: 106 Resp: 458678
Ion Ratio Lower Upper
106 100
91 194.8 155.3 232.9









#69
o-Xylene
Concen: 49.248 ug/l
RT: 10.69 min Scan# 1565
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion:106 Resp: 223652
Ion Ratio Lower Upper
106 100
91 204.3 102.7 308.1




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VX0






#70
Styrene
Concen: 50.714 ug/l
RT: 10.71 min Scan# 1568
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion:104 Resp: 386238
Ion Ratio Lower Upper
104 100
78 45.9 36.2 54.2
103 55.2 44.2 66.4



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VX0
Ion 103.00 (102.70 to 103.70): V





#71

Bromoform

Concen: 50.181 ug/l

RT: 10.85 min Scan# 1591

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:173 Resp: 112213

Ion Ratio Lower Upper

173 100

175 49.4 24.1 72.3




254 0.2 0.2 0.2



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23


Tgt Ion:152 Resp: 178222

Ion Ratio Lower Upper

152 100

115 76.6 38.6 115.7



150 171.7 0.0 347.4




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 49.535 ug/l

RT: 11.02 min Scan# 1618

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion:105 Resp: 598439

Ion Ratio Lower Upper




105 100

120 27.7 14.1 42.4



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#74

N-ethyl acetate

Concen: 51.287 ug/l

RT: 10.90 min Scan# 1598

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Tgt Ion: 43 Resp: 234164


Ion Ratio Lower Upper

43 100

70 44.7 36.6 54.8

55 23.8 18.7 28.1

61 25.4 20.6 31.0






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0





#75

1,1,2,2-Tetrachloroethane

Concen: 49.871 ug/l

RT: 11.27 min Scan# 1659

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 83 Resp: 199535

Ion Ratio Lower Upper

83 100

131 11.2 5.5 16.4




85 64.1 32.3 96.8



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): VX0



#76

1,2,3-Trichloropropane

Concen: 67.749 ug/l

RT: 11.29 min Scan# 1663

Delta R.T. 0.00 min

Lab File: VX010502.D


Acq: 27 Jun 2019 20:23

Tgt Ion: 75 Resp: 220026

Ion Ratio Lower Upper



75 100


77 32.1 22.6 67.8



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#77

Bromobenzene

Concen: 48.675 ug/l

RT: 11.26 min Scan# 1657

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:156 Resp: 164203

Ion Ratio Lower Upper

156 100

77 131.3 66.3 198.7




158 97.8 48.9 146.7



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VX0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 49.342 ug/l

RT: 11.36 min Scan# 1674

Delta R.T. 0.00 min

Lab File: VX010502.D


Acq: 27 Jun 2019 20:23

Tgt Ion: 91 Resp: 655149

Ion Ratio Lower Upper



91 100


120 24.7 12.4 37.4



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 120.00 (119.70 to 120.70): V









#79
2-Chlorotoluene
Concen: 48.920 ug/l
RT: 11.42 min Scan# 1684
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion: 91 Resp: 386793
Ion Ratio Lower Upper
91 100
126 37.5 18.7 56.1




Abundance Ion 91.00 (90.70 to 91.70): VX0
Ion 125.90 (125.60 to 126.60): V






#80
1,3,5-Trimethylbenzene
Concen: 49.377 ug/l
RT: 11.51 min Scan# 1698
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion: 105 Resp: 491777
Ion Ratio Lower Upper
105 100
120 50.6 25.2 75.6



Abundance Ion 105.00 (104.70 to 105.70): V
Ion 120.00 (119.70 to 120.70): V









#81
trans-1,4-Dichloro-2-butene
Concen: 44.838 ug/l
RT: 11.07 min Scan# 1627
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 75 Resp: 62420
Ion Ratio Lower Upper
75 100
53 86.8 66.1 99.1
89 47.9 37.8 56.6



#82
4-Chlorotoluene
Concen: 49.148 ug/l
RT: 11.51 min Scan# 1698
Delta R.T. -0.01 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion: 91 Resp: 442178
Ion Ratio Lower Upper
91 100
126 33.3 16.2 48.6









#83
 tert-Butylbenzene
 Concen: 49.471 ug/l
 RT: 11.77 min Scan# 1741
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion:119 Resp: 494524
 Ion Ratio Lower Upper
 119 100
 91 55.5 27.6 82.7
 134 23.5 11.8 35.4



#84
 1,2,4-Trimethylbenzene
 Concen: 50.375 ug/l
 RT: 11.80 min Scan# 1747
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Tgt Ion:105 Resp: 504060
 Ion Ratio Lower Upper
 105 100
 120 46.7 23.1 69.2





#85

sec-Butylbenzene

Concen: 49.947 ug/l

RT: 11.94 min Scan# 1770

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion:105 Resp: 586231

Ion Ratio Lower Upper




105 100

134 21.0 10.7 31.9



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 49.530 ug/l

RT: 12.06 min Scan# 1789

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23


Tgt Ion:119 Resp: 537709

Ion Ratio Lower Upper

119 100

134 27.4 13.7 41.1



91 22.4 11.3 33.8




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): VX0









#87
 1,3-Dichlorobenzene
 Concen: 48.534 ug/l
 RT: 12.02 min Scan# 1783
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion:146 Resp: 285793
 Ion Ratio Lower Upper
 146 100
 111 37.9 19.1 57.1
 148 63.8 32.5 97.4




Abundance Ion 145.90 (145.60 to 146.60): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 147.90 (147.60 to 148.60): V






#88
 1,4-Dichlorobenzene
 Concen: 47.703 ug/l
 RT: 12.10 min Scan# 1795
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Tgt Ion:146 Resp: 285969
 Ion Ratio Lower Upper
 146 100
 111 36.3 18.1 54.3
 148 64.0 32.1 96.3



Abundance Ion 145.90 (145.60 to 146.60): V
 Ion 110.90 (110.60 to 111.60): V
 Ion 147.90 (147.60 to 148.60): V









#89
n-Butylbenzene
Concen: 49.178 ug/l
RT: 12.38 min Scan# 1842
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 91 Resp: 453674
Ion Ratio Lower Upper
91 100
92 52.2 26.2 78.5
134 28.3 14.1 42.4



#90
Hexachloroethane
Concen: 49.635 ug/l
RT: 12.60 min Scan# 1877
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion: 117 Resp: 93174
Ion Ratio Lower Upper
117 100
201 88.1 44.0 131.8





#91

1,2-Dichlorobenzene

Concen: 48.203 ug/l

RT: 12.39 min Scan# 1843

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:146 Resp: 282107

Ion Ratio Lower Upper

146 100

111 38.9 19.2 57.6




148 64.3 32.0 96.0



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 47.940 ug/l

RT: 13.00 min Scan# 1943

Delta R.T. 0.00 min

Lab File: VX010502.D

Acq: 27 Jun 2019 20:23


Tgt Ion: 75 Resp: 41355

Ion Ratio Lower Upper

75 100

155 106.4 54.4 163.1



157 135.2 69.5 208.3




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V









#93
 1,2,4-Trichlorobenzene
 Concen: 49.174 ug/l
 RT: 13.65 min Scan# 2049
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion:180 Resp: 199577
 Ion Ratio Lower Upper
 180 100
 182 94.8 48.0 144.0
 145 29.6 14.4 43.4



#94
 Hexachlorobutadiene
 Concen: 48.597 ug/l
 RT: 13.78 min Scan# 2071
 Delta R.T. 0.00 min
 Lab File: VX010502.D
 Acq: 27 Jun 2019 20:23

Tgt Ion:225 Resp: 96396
 Ion Ratio Lower Upper
 225 100
 223 63.3 30.9 92.7
 227 64.6 31.9 95.5









#95
Naphthalene
Concen: 50.509 ug/l
RT: 13.83 min Scan# 2080
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion:128 Resp: 629968
Ion Ratio Lower Upper
128 100
127 12.8 10.2 15.4
129 10.9 8.9 13.3



#96
1,2,3-Trichlorobenzene
Concen: 49.494 ug/l
RT: 14.02 min Scan# 2110
Delta R.T. 0.00 min
Lab File: VX010502.D
Acq: 27 Jun 2019 20:23

Tgt Ion:180 Resp: 199722
Ion Ratio Lower Upper
180 100
182 96.4 47.8 143.3
145 30.5 15.2 45.6

