

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062719\
 Data File : VX010509.D
 Acq On : 28 Jun 2019 01:03
 Operator : JC/SP
 Sample : PB120955TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB120955TB

Quant Time: Jun 28 08:04:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	269260	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	411037	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	379139	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	173733	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	125509	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
35) Dibromofluoromethane	5.50	113	124080	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
50) Toluene-d8	8.71	98	491561	51.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.18%	
62) 4-Bromofluorobenzene	11.13	95	162157	46.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.36%	
Target Compounds						
20) Methylene Chloride	2.85	84	5490	2.108	ug/l	95
43) Isopropyl Acetate	6.45	43	12599	2.266	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062719\
Data File : VX010509.D
Acq On : 28 Jun 2019 01:03
Operator : JC/SP
Sample : PB120955TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB120955TB

Quant Time: Jun 28 08:04:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

