

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062719\
 Data File : VX010510.D
 Acq On : 28 Jun 2019 01:27
 Operator : JC/SP
 Sample : K3507-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NB-07-062619-A

Quant Time: Jun 28 08:07:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	248330	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	382517	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	355844	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	168294	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118637	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
35) Dibromofluoromethane	5.50	113	116633	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
50) Toluene-d8	8.71	98	458569	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
62) 4-Bromofluorobenzene	11.13	95	155108	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
Target Compounds						
20) Methylene Chloride	2.85	84	81464	33.924	ug/l	99
30) Chloroform	5.21	83	4274	1.076	ug/l	91
43) Isopropyl Acetate	6.45	43	10732	2.074	ug/l	96
85) sec-Butylbenzene	11.94	105	13309	1.201	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062719\
Data File : VX010510.D
Acq On : 28 Jun 2019 01:27
Operator : JC/SP
Sample : K3507-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NB-07-062619-A

Quant Time: Jun 28 08:07:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

