

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062719\
 Data File : VX010511.D
 Acq On : 28 Jun 2019 01:50
 Operator : JC/SP
 Sample : K3507-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NB-07-062619-B

Quant Time: Jun 28 08:08:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	251403	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	381482	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	357857	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	168816	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118301	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
35) Dibromofluoromethane	5.50	113	116265	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
50) Toluene-d8	8.71	98	458458	51.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.68%	
62) 4-Bromofluorobenzene	11.13	95	156615	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
Target Compounds						
16) Acetone	2.44	43	17287	14.633	ug/l	96
20) Methylene Chloride	2.86	84	34967	14.383	ug/l	95
43) Isopropyl Acetate	6.45	43	10765	2.086	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062719\
Data File : VX010511.D
Acq On : 28 Jun 2019 01:50
Operator : JC/SP
Sample : K3507-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NB-07-062619-B

Quant Time: Jun 28 08:08:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

