

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062719\
 Data File : VX010515.D
 Acq On : 28 Jun 2019 03:24
 Operator : JC/SP
 Sample : K3157-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-062619-B

Quant Time: Jun 28 08:15:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	260187	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	394192	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	369206	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	172901	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	123073	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
35) Dibromofluoromethane	5.50	113	119635	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
50) Toluene-d8	8.71	98	475049	51.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.96%	
62) 4-Bromofluorobenzene	11.13	95	159866	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
Target Compounds						
20) Methylene Chloride	2.86	84	8750	3.478	ug/l	94
43) Isopropyl Acetate	6.45	43	10302	1.932	ug/l	93
95) Naphthalene	13.83	128	31731	2.622	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062719\
Data File : VX010515.D
Acq On : 28 Jun 2019 03:24
Operator : JC/SP
Sample : K3157-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-02-062619-B

Quant Time: Jun 28 08:15:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

