

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062722\
 Data File : VX029814.D
 Acq On : 27 Jun 2022 12:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Jul 11 06:23:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062722W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 27 15:32:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	268659	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	531682	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	475995	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	207693	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	202774	48.718	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.440%
35) Dibromofluoromethane	5.391	113	169061	49.498	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.000%
50) Toluene-d8	8.653	98	668680	49.993	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.980%
62) 4-Bromofluorobenzene	11.079	95	247654	50.630	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.260%

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.166	85	183636	54.660	ug/l	97
3) Chloromethane	1.288	50	190361	50.894	ug/l	99
4) Vinyl Chloride	1.374	62	204524	51.878	ug/l	99
5) Bromomethane	1.611	94	40103	48.762	ug/l	94
6) Chloroethane	1.685	64	109000	50.410	ug/l	98
7) Trichlorofluoromethane	1.886	101	226926	50.576	ug/l	96
8) Diethyl Ether	2.136	74	103494	48.606	ug/l	79
9) 1,1,2-Trichlorotrifluo...	2.331	101	156102	49.032	ug/l	93
10) Methyl Iodide	2.453	142	197789	54.152	ug/l	96
11) Tert butyl alcohol	2.983	59	257478	251.767	ug/l	97
12) 1,1-Dichloroethene	2.319	96	164024	50.082	ug/l	97
13) Acrolein	2.239	56	170419	234.002	ug/l	96
14) Allyl chloride	2.666	41	308801	51.647	ug/l	92
15) Acrylonitrile	3.068	53	614461	250.815	ug/l	99
16) Acetone	2.386	43	473919	238.835	ug/l	96
17) Carbon Disulfide	2.514	76	441306	52.361	ug/l	99
18) Methyl Acetate	2.709	43	279492	46.975	ug/l	93
19) Methyl tert-butyl Ether	3.117	73	563646	49.755	ug/l	97
20) Methylene Chloride	2.788	84	196528	47.874	ug/l	95
21) trans-1,2-Dichloroethene	3.093	96	178088	50.794	ug/l	97
22) Diisopropyl ether	3.763	45	613907	50.211	ug/l	88
23) Vinyl Acetate	3.721	43	2708092	260.412	ug/l	# 92
24) 1,1-Dichloroethane	3.611	63	333645	50.188	ug/l	98
25) 2-Butanone	4.556	43	839054	249.004	ug/l	92
26) 2,2-Dichloropropane	4.477	77	235523	50.574	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	210649	49.181	ug/l	99
28) Bromochloromethane	4.903	49	120424	45.753	ug/l	86
29) Tetrahydrofuran	5.013	42	570352	253.885	ug/l	88
30) Chloroform	5.099	83	321838	49.775	ug/l	99
31) Cyclohexane	5.477	56	324428	52.327	ug/l	90
32) 1,1,1-Trichloroethane	5.397	97	257857	51.411	ug/l	96
36) 1,1-Dichloropropene	5.696	75	251339	50.239	ug/l	99
37) Ethyl Acetate	4.715	43	317145	51.447	ug/l	98
38) Carbon Tetrachloride	5.678	117	209033	51.506	ug/l	99
39) Methylcyclohexane	7.385	83	318373	52.384	ug/l	93
40) Benzene	6.044	78	789397	50.266	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062722\
 Data File : VX029814.D
 Acq On : 27 Jun 2022 12:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Quant Time: Jul 11 06:23:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062722W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 27 15:32:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	173761	49.183	ug/l	92
42) 1,2-Dichloroethane	6.086	62	251493	49.613	ug/l	98
43) Isopropyl Acetate	6.348	43	480185	50.914	ug/l	97
44) Trichloroethene	7.129	130	182749	47.432	ug/l	90
45) 1,2-Dichloropropane	7.434	63	200031	48.215	ug/l	98
46) Dibromomethane	7.586	93	129672	48.877	ug/l #	87
47) Bromodichloromethane	7.824	83	241041	50.854	ug/l	99
48) Methyl methacrylate	7.696	41	233396	49.749	ug/l	90
49) 1,4-Dioxane	7.683	88	107542	955.668	ug/l	96
51) 4-Methyl-2-Pentanone	8.580	43	1560849	251.054	ug/l	94
52) Toluene	8.720	92	468567	50.268	ug/l	96
53) t-1,3-Dichloropropene	8.982	75	270974	52.800	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	310891	51.817	ug/l #	91
55) 1,1,2-Trichloroethane	9.153	97	192746	49.091	ug/l	95
56) Ethyl methacrylate	9.122	69	327211	51.233	ug/l	89
57) 1,3-Dichloropropane	9.311	76	335636	50.671	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	774388	253.481	ug/l	93
59) 2-Hexanone	9.433	43	1233749	255.569	ug/l	92
60) Dibromochloromethane	9.525	129	170960	51.230	ug/l	98
61) 1,2-Dibromoethane	9.616	107	201422	51.570	ug/l	98
64) Tetrachloroethene	9.281	164	135892	49.480	ug/l	95
65) Chlorobenzene	10.079	112	477810	47.778	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	159831	49.546	ug/l	96
67) Ethyl Benzene	10.195	91	904411	49.866	ug/l	99
68) m/p-Xylenes	10.305	106	675303	98.993	ug/l	98
69) o-Xylene	10.646	106	327558	48.574	ug/l	95
70) Styrene	10.659	104	546912	49.304	ug/l	98
71) Bromoform	10.805	173	111569	51.740	ug/l #	99
73) Isopropylbenzene	10.963	105	861012	49.954	ug/l	98
74) N-amyl acetate	10.848	43	417402	50.466	ug/l	90
75) 1,1,2,2-Tetrachloroethane	11.213	83	325036	48.257	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	359111	60.993	ug/l	89
77) Bromobenzene	11.201	156	181837	48.041	ug/l	80
78) n-propylbenzene	11.305	91	1071784	51.384	ug/l	97
79) 2-Chlorotoluene	11.366	91	616367	49.056	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	703384	48.868	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	102348	52.870	ug/l	96
82) 4-Chlorotoluene	11.457	91	704414	49.215	ug/l	97
83) tert-Butylbenzene	11.719	119	697647	49.965	ug/l	92
84) 1,2,4-Trimethylbenzene	11.756	105	718545	49.499	ug/l	100
85) sec-Butylbenzene	11.890	105	913631	50.311	ug/l	97
86) p-Isopropyltoluene	12.012	119	738311	51.399	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	357646	49.190	ug/l	96
88) 1,4-Dichlorobenzene	12.042	146	357985	48.612	ug/l	96
89) n-Butylbenzene	12.335	91	697294	51.821	ug/l	97
90) Hexachloroethane	12.542	117	117891	54.285	ug/l	77
91) 1,2-Dichlorobenzene	12.335	146	349010	48.279	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	76943	50.152	ug/l	80
93) 1,2,4-Trichlorobenzene	13.591	180	207442	49.554	ug/l	96
94) Hexachlorobutadiene	13.725	225	72472	48.216	ug/l	96
95) Naphthalene	13.780	128	894152	50.444	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	210202	49.697	ug/l	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062722\
Data File : VX029814.D
Acq On : 27 Jun 2022 12:58
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Quant Time: Jul 11 06:23:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062722W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 27 15:32:24 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 11 06:23:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062722W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 27 15:32:24 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:168 Resp: 268659
Ion Ratio Lower Upper
168 100
99 63.5 41.7 62.5#



#2
Dichlorodifluoromethane
Concen: 54.660 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 85 Resp: 183636
Ion Ratio Lower Upper
85 100
87 30.8 16.2 48.6









#3
Chloromethane
Concen: 50.894 ug/l
RT: 1.288 min Scan# 33
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 50 Resp: 190361
Ion Ratio Lower Upper
50 100
52 33.8 26.4 39.6



#4
Vinyl Chloride
Concen: 51.878 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 62 Resp: 204524
Ion Ratio Lower Upper
62 100
64 31.6 25.5 38.3









#5
Bromomethane
Concen: 48.762 ug/l
RT: 1.611 min Scan# 86
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 94 Resp: 40103
Ion Ratio Lower Upper
94 100
96 87.9 75.3 112.9



#6
Chloroethane
Concen: 50.410 ug/l
RT: 1.685 min Scan# 98
Delta R.T. -0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 64 Resp: 109000
Ion Ratio Lower Upper
64 100
66 32.6 25.3 37.9









#7
Trichlorofluoromethane
Concen: 50.576 ug/l
RT: 1.886 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:101 Resp: 226926
Ion Ratio Lower Upper
101 100
103 65.9 50.5 75.7



#8
Diethyl Ether
Concen: 48.606 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 74 Resp: 103494
Ion Ratio Lower Upper
74 100
45 102.8 42.0 126.0









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.032 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:101 Resp: 156102
Ion Ratio Lower Upper
101 100
85 43.8 33.5 50.3
151 69.3 61.8 92.6



#10
Methyl Iodide
Concen: 54.152 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion:142 Resp: 197789
Ion Ratio Lower Upper
142 100
127 39.5 33.9 50.9
141 14.0 11.4 17.2









#11
Tert butyl alcohol
Concen: 251.767 ug/l
RT: 2.983 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 59 Resp: 257478
Ion Ratio Lower Upper
59 100
57 9.4 8.3 12.5



#12
1,1-Dichloroethene
Concen: 50.082 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 96 Resp: 164024
Ion Ratio Lower Upper
96 100
61 152.6 124.6 187.0
98 61.3 51.0 76.4









#13
Acrolein
Concen: 234.002 ug/l
RT: 2.239 min Scan# 189
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 170419
Ion Ratio Lower Upper
56 100
55 71.6 55.0 82.6



#14
Allyl chloride
Concen: 51.647 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 41 Resp: 308801
Ion Ratio Lower Upper
41 100
39 58.3 50.4 75.6
76 33.7 32.7 49.1









#15
Acrylonitrile
Concen: 250.815 ug/l
RT: 3.068 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 53 Resp: 614461
Ion Ratio Lower Upper
53 100
52 82.6 65.7 98.5
51 34.3 28.4 42.6



#16
Acetone
Concen: 238.835 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 43 Resp: 473919
Ion Ratio Lower Upper
43 100
58 33.4 28.8 43.2









#17
Carbon Disulfide
Concen: 52.361 ug/l
RT: 2.514 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 76 Resp: 441306
Ion Ratio Lower Upper
76 100
78 9.3 7.2 10.8



#18
Methyl Acetate
Concen: 46.975 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 43 Resp: 279492
Ion Ratio Lower Upper
43 100
74 25.0 22.8 34.2









#19
Methyl tert-butyl Ether
Concen: 49.755 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 73 Resp: 563646
Ion Ratio Lower Upper
73 100
57 22.6 16.8 25.2



#20
Methylene Chloride
Concen: 47.874 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 84 Resp: 196528
Ion Ratio Lower Upper
84 100
49 115.3 85.5 128.3
51 35.1 26.6 39.8
86 65.5 51.5 77.3









#21
trans-1,2-Dichloroethene
Concen: 50.794 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 96 Resp: 178088
Ion Ratio Lower Upper
96 100
61 136.4 105.4 158.2
98 62.9 50.0 75.0



#22
Diisopropyl ether
Concen: 50.211 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 45 Resp: 613907
Ion Ratio Lower Upper
45 100
43 75.5 50.8 76.2
87 28.5 26.6 39.8
59 12.0 10.5 15.7









#23
 Vinyl Acetate
 Concen: 260.412 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: VX029814.D
 Acq: 27 Jun 2022 12:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 43 Resp: 2708092
 Ion Ratio Lower Upper
 43 100
 86 10.4 10.8 16.2#



#24
 1,1-Dichloroethane
 Concen: 50.188 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX029814.D
 Acq: 27 Jun 2022 12:58

Tgt Ion: 63 Resp: 333645
 Ion Ratio Lower Upper
 63 100
 98 6.9 3.9 11.7
 100 4.3 2.4 7.2









#25
2-Butanone
Concen: 249.004 ug/l
RT: 4.556 min Scan# 569
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 839054
Ion Ratio Lower Upper
43 100
72 26.0 24.4 36.6



#26
2,2-Dichloropropane
Concen: 50.574 ug/l
RT: 4.477 min Scan# 556
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 77 Resp: 235523
Ion Ratio Lower Upper
77 100
97 24.0 11.8 35.4





#27

cis-1,2-Dichloroethene

Concen: 49.181 ug/l

RT: 4.489 min Scan# 558

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 96 Resp: 210649

Ion Ratio Lower Upper

96 100

61 140.4 0.0 283.8

98 64.1 0.0 129.8



#28

Bromochloromethane

Concen: 45.753 ug/l

RT: 4.903 min Scan# 626

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58




Tgt Ion: 49 Resp: 120424


Ion Ratio Lower Upper

49 100

129 2.0 0.0 4.4

130 75.0 70.7 106.1





#29

Tetrahydrofuran

Concen: 253.885 ug/l

RT: 5.013 min Scan# 644

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 42 Resp: 570352

Ion Ratio Lower Upper

42 100

72 46.2 44.1 66.1

71 43.1 40.5 60.7



#30

Chloroform

Concen: 49.775 ug/l

RT: 5.099 min Scan# 658

Delta R.T. 0.000 min

Lab File: VX029814.D




Acq: 27 Jun 2022 12:58


Tgt Ion: 83 Resp: 321838

Ion Ratio Lower Upper

83 100

85 65.0 52.7 79.1









#31
Cyclohexane
Concen: 52.327 ug/l
RT: 5.477 min Scan# 720
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 324428
Ion Ratio Lower Upper
56 100
69 29.6 26.7 40.1
84 85.0 77.2 115.8



#32
1,1,1-Trichloroethane
Concen: 51.411 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 97 Resp: 257857
Ion Ratio Lower Upper
97 100
99 62.5 51.8 77.6
61 45.7 34.0 51.0





#33

1,2-Dichloroethane-d4

Concen: 48.718 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 65 Resp: 202774

Ion Ratio Lower Upper

65 100

67 55.7 0.0 107.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58




Tgt Ion: 114 Resp: 531682


Ion Ratio Lower Upper

114 100

63 20.1 0.0 36.8

88 16.2 0.0 32.4





#35

Dibromofluoromethane

Concen: 49.498 ug/l

RT: 5.391 min Scan# 706

Delta R.T. 0.000 min

Lab File: VX029814.D


Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050






Tgt Ion:113 Resp: 169061

Ion Ratio Lower Upper

113 100

111 104.5 82.8 124.2

192 16.2 16.9 25.3#



#36

1,1-Dichloropropene

Concen: 50.239 ug/l

RT: 5.696 min Scan# 756

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58




Tgt Ion: 75 Resp: 251339


Ion Ratio Lower Upper

75 100

110 35.9 18.3 54.8

77 31.1 24.8 37.2









#37
Ethyl Acetate
Concen: 51.447 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 317145
Ion Ratio Lower Upper
43 100
61 15.3 12.6 19.0
70 11.9 10.2 15.2



#38
Carbon Tetrachloride
Concen: 51.506 ug/l
RT: 5.678 min Scan# 753
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 117 Resp: 209033
Ion Ratio Lower Upper
117 100
119 95.7 77.3 115.9
121 31.3 24.2 36.2









#39
Methylcyclohexane
Concen: 52.384 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 318373
Ion Ratio Lower Upper
83 100
55 75.3 54.6 81.8
98 44.6 37.4 56.2



#40
Benzene
Concen: 50.266 ug/l
RT: 6.044 min Scan# 813
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 78 Resp: 789397
Ion Ratio Lower Upper
78 100
77 22.7 19.3 28.9









#41
Methacrylonitrile
Concen: 49.183 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 41 Resp: 173761
Ion Ratio Lower Upper
41 100
39 52.5 43.0 64.4
67 77.4 70.2 105.4
52 27.1 26.0 39.0



#42
1,2-Dichloroethane
Concen: 49.613 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 62 Resp: 251493
Ion Ratio Lower Upper
62 100
98 9.5 0.0 20.4









#43
Isopropyl Acetate
Concen: 50.914 ug/l
RT: 6.348 min Scan# 863
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 480185
Ion Ratio Lower Upper
43 100
61 22.2 18.5 27.7
87 13.9 13.0 19.6



#44
Trichloroethene
Concen: 47.432 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion:130 Resp: 182749
Ion Ratio Lower Upper
130 100
95 104.4 0.0 188.8





#45

1,2-Dichloropropane

Concen: 48.215 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 63 Resp: 200031

Ion Ratio Lower Upper

63 100

65 33.2 25.8 38.6



#46

Dibromomethane

Concen: 48.877 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58




Tgt Ion: 93 Resp: 129672


Ion Ratio Lower Upper

93 100

95 83.5 65.8 98.6

174 83.8 85.4 128.2#









#47
Bromodichloromethane
Concen: 50.854 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 83 Resp: 241041
Ion Ratio Lower Upper
83 100
85 64.8 52.2 78.2
127 7.7 6.8 10.2



#48
Methyl methacrylate
Concen: 49.749 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 41 Resp: 233396
Ion Ratio Lower Upper
41 100
69 89.8 82.5 123.7
39 51.1 43.8 65.8





#49

1,4-Dioxane

Concen: 955.668 ug/l

RT: 7.683 min Scan# 1082

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 88 Resp: 107542

Ion Ratio Lower Upper

88 100

43 28.6 21.9 32.9

58 70.3 53.7 80.5



#50

Toluene-d8

Concen: 49.993 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. 0.000 min

Lab File: VX029814.D




Acq: 27 Jun 2022 12:58


Tgt Ion: 98 Resp: 668680

Ion Ratio Lower Upper

98 100

100 65.3 52.6 78.8





#51

4-Methyl-2-Pentanone

Concen: 251.054 ug/l

RT: 8.580 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 43 Resp: 1560849

Ion Ratio Lower Upper

43 100

58 41.1 36.1 54.1



#52

Toluene

Concen: 50.268 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. 0.000 min

Lab File: VX029814.D




Acq: 27 Jun 2022 12:58


Tgt Ion: 92 Resp: 468567

Ion Ratio Lower Upper

92 100

91 176.9 137.5 206.3









#53
 t-1,3-Dichloropropene
 Concen: 52.800 ug/l
 RT: 8.982 min Scan# 1295
 Delta R.T. 0.000 min
 Lab File: VX029814.D
 Acq: 27 Jun 2022 12:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 75 Resp: 270974
 Ion Ratio Lower Upper
 75 100
 77 32.1 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 51.817 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. 0.000 min
 Lab File: VX029814.D
 Acq: 27 Jun 2022 12:58

Tgt Ion: 75 Resp: 310891
 Ion Ratio Lower Upper
 75 100
 77 33.0 24.8 37.2
 39 44.9 29.6 44.4#





#55

1,1,2-Trichloroethane

Concen: 49.091 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 97 Resp: 192746





Ion Ratio Lower Upper

97 100

83 90.1 66.2 99.4

85 58.7 42.9 64.3

99 61.2 49.0 73.6



#56

Ethyl methacrylate

Concen: 51.233 ug/l

RT: 9.122 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58




Tgt Ion: 69 Resp: 327211


Ion Ratio Lower Upper

69 100

41 64.4 43.8 65.6

39 30.3 22.2 33.2





#57

1,3-Dichloropropane

Concen: 50.671 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 76 Resp: 335636

Ion Ratio Lower Upper

76 100

78 32.8 25.8 38.8



#58

2-Chloroethyl Vinyl ether

Concen: 253.481 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX029814.D




Acq: 27 Jun 2022 12:58


Tgt Ion: 63 Resp: 774388

Ion Ratio Lower Upper

63 100

106 26.3 23.9 35.9









#59
2-Hexanone
Concen: 255.569 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 1233749
Ion Ratio Lower Upper
43 100
58 57.2 31.9 95.7



#60
Dibromochloromethane
Concen: 51.230 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 129 Resp: 170960
Ion Ratio Lower Upper
129 100
127 78.5 40.3 120.8









#61
1,2-Dibromoethane
Concen: 51.570 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:107 Resp: 201422
Ion Ratio Lower Upper
107 100
109 95.0 74.4 111.6



#62
4-Bromofluorobenzene
Concen: 50.630 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 95 Resp: 247654
Ion Ratio Lower Upper
95 100
174 61.7 0.0 163.0
176 57.9 0.0 155.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:117 Resp: 475995

Ion Ratio Lower Upper

117 100

82 66.4 41.8 62.8#

119 32.5 24.8 37.2



#64

Tetrachloroethene

Concen: 49.480 ug/l

RT: 9.281 min Scan# 1344

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Tgt Ion:164 Resp: 135892




Ion Ratio Lower Upper


164 100

166 117.8 101.2 151.8

129 95.4 73.0 109.4

131 94.1 74.2 111.2









#65
Chlorobenzene
Concen: 47.778 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:112 Resp: 477810
Ion Ratio Lower Upper
112 100
114 32.1 25.8 38.8



#66
1,1,1,2-Tetrachloroethane
Concen: 49.546 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion:131 Resp: 159831
Ion Ratio Lower Upper
131 100
133 95.9 48.3 144.8
119 72.3 32.6 98.0





#67

Ethyl Benzene

Concen: 49.866 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion: 91 Resp: 904411

Ion Ratio Lower Upper

91 100

106 31.3 24.6 36.8



#68

m/p-Xylenes

Concen: 98.993 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. 0.000 min

Lab File: VX029814.D




Acq: 27 Jun 2022 12:58


Tgt Ion: 106 Resp: 675303

Ion Ratio Lower Upper

106 100

91 203.6 160.4 240.6





#69

o-Xylene

Concen: 48.574 ug/l

RT: 10.646 min Scan# 1568

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:106 Resp: 327558

Ion Ratio Lower Upper

106 100

91 219.4 106.1 318.1



#70

Styrene

Concen: 49.304 ug/l

RT: 10.659 min Scan# 1570

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58




Tgt Ion:104 Resp: 546912


Ion Ratio Lower Upper

104 100

78 51.5 38.6 57.8

103 54.7 43.8 65.8





#71

Bromoform

Concen: 51.740 ug/l

RT: 10.805 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:173 Resp: 111569

Ion Ratio Lower Upper

173 100

175 48.6 24.1 72.2

254 0.0 0.1 0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58




Tgt Ion:152 Resp: 207693


Ion Ratio Lower Upper

152 100

115 91.5 37.5 112.7

150 176.9 0.0 338.2





#73

Isopropylbenzene

Concen: 49.954 ug/l

RT: 10.963 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Tgt Ion:105 Resp: 861012

Ion Ratio Lower Upper

105 100

120 26.0 13.5 40.5



#74

N-ethyl acetate

Concen: 50.466 ug/l

RT: 10.848 min Scan# 1601

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Tgt Ion: 43 Resp: 417402




Ion Ratio Lower Upper


43 100

70 43.6 42.5 63.7

55 24.9 23.1 34.7

61 26.0 22.8 34.2





#75

1,1,2,2-Tetrachloroethane

Concen: 48.257 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion: 83 Resp: 325036

Ion Ratio Lower Upper

83 100

131 8.6 5.1 15.3

85 64.3 32.6 97.7



#76

1,2,3-Trichloropropane

Concen: 60.993 ug/l

RT: 11.244 min Scan# 1666

Delta R.T. 0.000 min

Lab File: VX029814.D




Acq: 27 Jun 2022 12:58


Tgt Ion: 75 Resp: 359111

Ion Ratio Lower Upper

75 100

77 31.9 19.4 58.2









#77
Bromobenzene
Concen: 48.041 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:156 Resp: 181837
Ion Ratio Lower Upper
156 100
77 195.3 78.5 235.3
158 94.8 50.2 150.7



#78
n-propylbenzene
Concen: 51.384 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 91 Resp: 1071784
Ion Ratio Lower Upper
91 100
120 21.9 11.6 34.8









#79
2-Chlorotoluene
Concen: 49.056 ug/l
RT: 11.366 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 616367
Ion Ratio Lower Upper
91 100
126 31.9 17.3 51.9



#80
1,3,5-Trimethylbenzene
Concen: 48.868 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 105 Resp: 703384
Ion Ratio Lower Upper
105 100
120 48.6 24.3 72.9









#81
trans-1,4-Dichloro-2-butene
Concen: 52.870 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 75 Resp: 102348
Ion Ratio Lower Upper
75 100
53 89.6 69.0 103.4
89 42.4 37.0 55.6



#82
4-Chlorotoluene
Concen: 49.215 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 91 Resp: 704414
Ion Ratio Lower Upper
91 100
126 28.2 14.9 44.5





#83

tert-Butylbenzene

Concen: 49.965 ug/l

RT: 11.719 min Scan# 1744

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:119 Resp: 697647

Ion Ratio Lower Upper

119 100

91 61.7 26.9 80.6

134 22.0 11.4 34.2



#84

1,2,4-Trimethylbenzene

Concen: 49.499 ug/l

RT: 11.756 min Scan# 1750

Delta R.T. 0.000 min

Lab File: VX029814.D




Acq: 27 Jun 2022 12:58


Tgt Ion:105 Resp: 718545

Ion Ratio Lower Upper

105 100

120 45.3 22.6 67.8









#85
 sec-Butylbenzene
 Concen: 50.311 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX029814.D
 Acq: 27 Jun 2022 12:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050


Tgt Ion:105 Resp: 913631
 Ion Ratio Lower Upper
 105 100
 134 19.0 10.2 30.4



#86
 p-Isopropyltoluene
 Concen: 51.399 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX029814.D
 Acq: 27 Jun 2022 12:58

Tgt Ion:119 Resp: 738311
 Ion Ratio Lower Upper
 119 100
 134 25.7 13.0 39.0
 91 24.9 11.3 33.9





#87

1,3-Dichlorobenzene

Concen: 49.190 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:146 Resp: 357646

Ion Ratio Lower Upper

146 100

111 45.2 19.7 59.1

148 62.7 31.6 94.7



#88

1,4-Dichlorobenzene

Concen: 48.612 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58




Tgt Ion:146 Resp: 357985


Ion Ratio Lower Upper

146 100

111 43.4 19.2 57.6

148 63.5 32.3 96.9









#89
n-Butylbenzene
Concen: 51.821 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 697294
Ion Ratio Lower Upper
91 100
92 53.9 27.8 83.3
134 23.5 13.3 39.8



#90
Hexachloroethane
Concen: 54.285 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 117 Resp: 117891
Ion Ratio Lower Upper
117 100
201 58.7 39.5 118.5









#91
1,2-Dichlorobenzene
Concen: 48.279 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050


Tgt Ion:146 Resp: 349010
Ion Ratio Lower Upper
146 100
111 47.7 21.0 63.0
148 63.6 32.6 97.7



#92
1,2-Dibromo-3-Chloropropane
Concen: 50.152 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.000 min
Lab File: VX029814.D
Acq: 27 Jun 2022 12:58

Tgt Ion: 75 Resp: 76943
Ion Ratio Lower Upper
75 100
155 75.3 47.0 141.2
157 93.6 57.9 173.6





#93

1,2,4-Trichlorobenzene

Concen: 49.554 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:180 Resp: 207442

Ion Ratio Lower Upper

180 100

182 92.2 47.7 143.1

145 34.3 15.2 45.5



#94

Hexachlorobutadiene

Concen: 48.216 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58




Tgt Ion:225 Resp: 72472


Ion Ratio Lower Upper

225 100

223 64.0 34.2 102.5

227 65.2 33.9 101.7





#95

Naphthalene

Concen: 50.444 ug/l

RT: 13.780 min Scan# 2082

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:128 Resp: 894152

Ion Ratio Lower Upper

128 100

127 12.8 10.2 15.4

129 11.1 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 49.697 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.000 min

Lab File: VX029814.D

Acq: 27 Jun 2022 12:58

Tgt Ion:180 Resp: 210202

Ion Ratio Lower Upper

180 100

182 92.9 50.0 149.8

145 37.2 16.2 48.5

