

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062722\
 Data File : VX029822.D
 Acq On : 27 Jun 2022 17:41
 Operator : JC/MD
 Sample : N3473-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-MW17411-TB-20220621RE

Quant Time: Jun 27 18:32:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062722W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 27 15:32:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	198011	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	423909	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	358582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	141475	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	152883	49.836	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.680%
35) Dibromofluoromethane	5.391	113	129841	47.680	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.360%
50) Toluene-d8	8.653	98	517807	48.556	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.120%
62) 4-Bromofluorobenzene	11.085	95	169597	43.487	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.980%
Target Compounds						
16) Acetone	2.386	43	75460	51.597	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062722\
Data File : VX029822.D
Acq On : 27 Jun 2022 17:41
Operator : JC/MD
Sample : N3473-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-MW174I1-TB-20220621RE

Quant Time: Jun 27 18:32:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062722W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 27 15:32:24 2022
Response via : Initial Calibration

