

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062722\
 Data File : VX029824.D
 Acq On : 27 Jun 2022 18:28
 Operator : JC/MD
 Sample : N3366-07DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-27DL

Quant Time: Jun 28 10:15:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062722W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 27 15:32:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	213416	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	456839	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	413665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	160015	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	168380	50.926	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.860%
35) Dibromofluoromethane	5.385	113	141141	48.094	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.180%
50) Toluene-d8	8.653	98	572071	49.777	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.560%
62) 4-Bromofluorobenzene	11.085	95	189968	45.199	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.400%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.977	59	85688	95.647	ug/l #	94
19) Methyl tert-butyl Ether	3.117	73	7796	0.866	ug/l #	89
22) Diisopropyl ether	3.770	45	7385	0.760	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062722\
Data File : VX029824.D
Acq On : 27 Jun 2022 18:28
Operator : JC/MD
Sample : N3366-07DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-27DL

Quant Time: Jun 28 10:15:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062722W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 27 15:32:24 2022
Response via : Initial Calibration

