



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062722\
 Data File : VX029809.D
 Acq On : 27 Jun 2022 10:26
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062722W.M
 Title : SW846 8260
 Last Update : Mon Jun 27 15:32:24 2022



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	13121	PASS
75	95	30	60	46.0	37488	PASS
95	95	100	100	100.0	81488	PASS
96	95	5	9	6.5	5317	PASS
173	174	0.00	2	1.0	549	PASS
174	95	50	100	65.0	52979	PASS
175	174	5	9	7.1	3740	PASS
176	174	95	101	95.9	50824	PASS
177	176	5	9	6.6	3350	PASS