

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062723\
 Data File : VX036350.D
 Acq On : 27 Jun 2023 14:50
 Operator : JC/MD
 Sample : 03289-12
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-05-0-2.0

Quant Time: Jun 28 04:20:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	528952	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	916333	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	815073	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	390193	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	413354	54.517	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.040%
35) Dibromofluoromethane	5.385	113	307992	50.975	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.940%
50) Toluene-d8	8.647	98	1164630	51.608	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.220%
62) 4-Bromofluorobenzene	11.085	95	424367	53.387	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062723\
Data File : VX036350.D
Acq On : 27 Jun 2023 14:50
Operator : JC/MD
Sample : 03289-12
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-05-0-2.0

Quant Time: Jun 28 04:20:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

