

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062723\
 Data File : VX036352.D
 Acq On : 27 Jun 2023 16:00
 Operator : JC/MD
 Sample : 03402-05RE
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP3-0623RE

Quant Time: Jun 28 04:21:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	311350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	541512	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	486427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	229552	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	398881	89.375	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	178.760%#
35) Dibromofluoromethane	5.385	113	290064	81.237	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	162.480%#
50) Toluene-d8	8.647	98	1108596	83.128	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	166.260%#
62) 4-Bromofluorobenzene	11.085	95	407352	86.718	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	173.440%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062723\
Data File : VX036352.D
Acq On : 27 Jun 2023 16:00
Operator : JC/MD
Sample : 03402-05RE
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP3-0623RE

Quant Time: Jun 28 04:21:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

