

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062723\
 Data File : VX036353.D
 Acq On : 27 Jun 2023 16:23
 Operator : JC/MD
 Sample : 03289-12
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-05-0-2.0

Quant Time: Jun 28 04:22:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	416238	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	713791	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	641718	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	310253	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	384925	64.515	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	129.020%#
35) Dibromofluoromethane	5.385	113	280743	59.650	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	119.300%
50) Toluene-d8	8.647	98	1092293	62.137	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	124.280%#
62) 4-Bromofluorobenzene	11.085	95	403430	65.155	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	130.300%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062723\
Data File : VX036353.D
Acq On : 27 Jun 2023 16:23
Operator : JC/MD
Sample : 03289-12
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-05-0-2.0

Quant Time: Jun 28 04:22:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

