

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062724\
 Data File : VX042075.D
 Acq On : 27 Jun 2024 15:25
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 28 04:05:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	28967	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	148898	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	133405	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	56616	28.392	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	94.633%
60) 4-Bromofluorobenzene	11.079	95	57555	27.872	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.900%
63) Toluene-d8	8.647	98	166989	29.258	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.533%
Target Compounds						
					Qvalue	
15) Acetone	2.386	58	973	3.175	ug/l	69
18) Methylene Chloride	2.782	84	1252	0.617	ug/l	91
62) Toluene	8.726	91	6450	0.955	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062724\
Data File : VX042075.D
Acq On : 27 Jun 2024 15:25
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 28 04:05:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 15 04:02:59 2024
Response via : Initial Calibration

