

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062724\
 Data File : VX042077.D
 Acq On : 27 Jun 2024 16:11
 Operator : JC/MD
 Sample : P3067-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001-WILLETS-PT-BLVD(JUNE)DL

Quant Time: Jun 28 04:05:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	29988	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	162063	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	147875	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	56697	27.465	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery =	91.533%		
60) 4-Bromofluorobenzene	11.079	95	63541	27.759	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery =	92.533%		
63) Toluene-d8	8.647	98	183354	28.981	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery =	96.600%		
Target Compounds						
					Qvalue	
15) Acetone	2.386	58	64159	202.234	ug/l	92
18) Methylene Chloride	2.794	84	1887	0.898	ug/l #	83
30) 2-Butanone	4.580	43	17745	12.556	ug/l	96
58) 4-Methyl-2-Pentanone	8.586	43	1440	0.543	ug/l	98
62) Toluene	8.720	91	226115	30.217	ug/l	99
82) p-Isopropyltoluene	12.012	119	2428	0.324	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062724\
Data File : VX042077.D
Acq On : 27 Jun 2024 16:11
Operator : JC/MD
Sample : P3067-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001-WILLETS-PT-BLVD(JUNE)DL

Quant Time: Jun 28 04:05:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 15 04:02:59 2024
Response via : Initial Calibration

