


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062724\
Data File : VX042065.D
Acq On : 27 Jun 2024 11:07
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Sat Jun 15 04:02:59 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.4	6474	PASS
75	95	30	60	46.0	19371	PASS
95	95	100	100	100.0	42069	PASS
96	95	5	9	7.0	2950	PASS
173	174	0.00	2	0.2	67	PASS
174	95	50	100	95.3	40104	PASS
175	174	5	9	7.4	2980	PASS
176	174	95	101	96.6	38736	PASS
177	176	5	9	6.8	2644	PASS