

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062818\
 Data File : VX002985.D
 Acq On : 28 Jun 2018 15:14
 Operator : JC/MD
 Sample : J3704-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE BLANK-WATERREF-4

Quant Time: Jun 29 09:52:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	257934	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	359505	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	327772	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	186019	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	172736	48.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.90%	
35) Dibromofluoromethane	5.51	113	139500	49.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.00%	
50) Toluene-d8	8.72	98	503447	48.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.86%	
62) 4-Bromofluorobenzene	11.14	95	174874	43.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.24%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062818\
Data File : VX002985.D
Acq On : 28 Jun 2018 15:14
Operator : JC/MD
Sample : J3704-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGEBLANK-WATERREF-4

Quant Time: Jun 29 09:52:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

