

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062818\
 Data File : VX002992.D
 Acq On : 28 Jun 2018 18:20
 Operator : JC/MD
 Sample : J3717-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 N48971

Quant Time: Jun 29 09:55:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	262449	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	358712	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	338500	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	197075	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	171670	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
35) Dibromofluoromethane	5.51	113	141257	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
50) Toluene-d8	8.72	98	509709	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
62) 4-Bromofluorobenzene	11.14	95	184135	46.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.06%	
Target Compounds						
16) Acetone	2.45	43	42628	28.38	ug/l	100
17) Carbon Disulfide	2.57	76	5113	0.74	ug/l	100
25) 2-Butanone	4.72	43	12383	5.55	ug/l	94
29) Tetrahydrofuran	5.17	42	36909	25.82	ug/l	98
84) 1,2,4-Trimethylbenzene	11.81	105	5391	0.47	ug/l	99
86) p-Isopropyltoluene	12.07	119	12265	1.02	ug/l	94
95) Naphthalene	13.84	128	11033	0.82	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062818\
Data File : VX002992.D
Acq On : 28 Jun 2018 18:20
Operator : JC/MD
Sample : J3717-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
N48971

Quant Time: Jun 29 09:55:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

