

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062818\
 Data File : VX002994.D
 Acq On : 28 Jun 2018 19:13
 Operator : JC/MD
 Sample : J3717-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1332

Quant Time: Jun 29 09:56:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	266958	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	374585	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	349917	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	212260	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	177547	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
35) Dibromofluoromethane	5.51	113	142956	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
50) Toluene-d8	8.72	98	528662	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	11.14	95	197483	47.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.54%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
29) Tetrahydrofuran	5.17	42	13040	8.97	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062818\
Data File : VX002994.D
Acq On : 28 Jun 2018 19:13
Operator : JC/MD
Sample : J3717-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1332

Quant Time: Jun 29 09:56:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

