

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062818\
 Data File : VX002995.D
 Acq On : 28 Jun 2018 19:40
 Operator : JC/MD
 Sample : J3717-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 N48966

Quant Time: Jun 29 09:57:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	272662	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	382896	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	357745	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	212671	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	181314	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
35) Dibromofluoromethane	5.51	113	147692	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
50) Toluene-d8	8.72	98	534202	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
62) 4-Bromofluorobenzene	11.14	95	198528	46.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.98%	
Target Compounds						
16) Acetone	2.45	43	34428	22.06	ug/l	Qvalue 100
25) 2-Butanone	4.71	43	7621	3.29	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062818\
Data File : VX002995.D
Acq On : 28 Jun 2018 19:40
Operator : JC/MD
Sample : J3717-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
N48966

Quant Time: Jun 29 09:57:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

