

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062818\
 Data File : VX003013.D
 Acq On : 29 Jun 2018 04:56
 Operator : JC/MD
 Sample : J3695-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LIC-FIELDBLANK-1-06-20-18

Quant Time: Jun 29 08:13:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	279903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	390425	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	364128	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	208319	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	184080	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
35) Dibromofluoromethane	5.51	113	150876	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
50) Toluene-d8	8.72	98	548730	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
62) 4-Bromofluorobenzene	11.14	95	194355	44.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.28%	
Target Compounds						
10) Methyl Iodide	2.52	142	1990	6.78	ug/l #	45
16) Acetone	2.45	43	19099	11.92	ug/l #	89
20) Methylene Chloride	2.86	84	5212	1.67	ug/l	88
25) 2-Butanone	4.71	43	4756	2.00	ug/l	98
51) 4-Methyl-2-Pentanone	8.65	43	1833	0.38	ug/l	94
65) Chlorobenzene	10.14	112	2329	0.25	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062818\
Data File : VX003013.D
Acq On : 29 Jun 2018 04:56
Operator : JC/MD
Sample : J3695-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LIC-FIELDBLANK-1-06-20-18

Quant Time: Jun 29 08:13:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

