

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062818\
 Data File : VX003014.D
 Acq On : 29 Jun 2018 05:22
 Operator : JC/MD
 Sample : J3720-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-13

Quant Time: Jun 29 10:18:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	289564	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	407510	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	376787	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	210554	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	193855	48.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.88%	
35) Dibromofluoromethane	5.51	113	156720	48.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.12%	
50) Toluene-d8	8.72	98	568728	48.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.52%	
62) 4-Bromofluorobenzene	11.14	95	200289	44.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062818\
Data File : VX003014.D
Acq On : 29 Jun 2018 05:22
Operator : JC/MD
Sample : J3720-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-13

Quant Time: Jun 29 10:18:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

