

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062819\
 Data File : VX010526.D
 Acq On : 28 Jun 2019 10:51
 Operator : JC/SP
 Sample : VX0628WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0628WBL01

Quant Time: Jun 29 04:52:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	276082	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	413863	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	363369	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	161626	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	137858	53.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.52%	
35) Dibromofluoromethane	5.49	113	137635	54.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.70%	
50) Toluene-d8	8.71	98	514924	53.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.30%	
62) 4-Bromofluorobenzene	11.13	95	164865	47.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.28%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062819\
Data File : VX010526.D
Acq On : 28 Jun 2019 10:51
Operator : JC/SP
Sample : VX0628WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0628WBL01

Quant Time: Jun 29 04:52:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

