

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062819\
 Data File : VX010542.D
 Acq On : 28 Jun 2019 18:08
 Operator : JC/SP
 Sample : K3565-13DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0577DL

Manual Integrations
 APPROVED

MMDadoda
 7/1/2019 11:36:08 AM

Quant Time: Jun 29 06:18:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	238853	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	369485	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	334826	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	158086	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115215	51.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.90%	
35) Dibromofluoromethane	5.50	113	112707	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
50) Toluene-d8	8.71	98	440823	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
62) 4-Bromofluorobenzene	11.13	95	146645	46.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.92%	
Target Compounds						
					Qvalue	
31) Cyclohexane	5.59	56	10843	3.352	ug/l	97
39) Methylcyclohexane	7.46	83	16792	4.351	ug/l	97
67) Ethyl Benzene	10.25	91	165836	14.451	ug/l	99
68) m/p-Xylenes	10.36	106	181368	40.788	ug/l	98
73) Isopropylbenzene	11.02	105	32950	3.075	ug/l	100
78) n-propylbenzene	11.36	91	45385	3.853	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	133736	15.138	ug/l	100
84) 1,2,4-Trimethylbenzene	11.80	105	254924	28.722	ug/l	99
85) sec-Butylbenzene	11.94	105	5698m	0.547	ug/l	
86) p-Isopropyltoluene	12.06	119	9789	1.017	ug/l	91
95) Naphthalene	13.83	128	93385	8.441	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062819\
Data File : VX010542.D
Acq On : 28 Jun 2019 18:08
Operator : JC/SP
Sample : K3565-13DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0577DL

Manual Integrations
APPROVED

MMDadoda
7/1/2019 11:36:08 AM

Quant Time: Jun 29 06:18:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

