

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062819\
 Data File : VX010545.D
 Acq On : 28 Jun 2019 19:18
 Operator : JC/SP
 Sample : K3572-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ERM-7

Quant Time: Jun 29 06:27:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	258818	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	400981	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	362685	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	165817	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	125762	51.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.64%	
35) Dibromofluoromethane	5.50	113	120104	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
50) Toluene-d8	8.71	98	470446	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
62) 4-Bromofluorobenzene	11.13	95	155748	45.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.92%	
Target Compounds						
16) Acetone	2.45	43	5392	4.434	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062819\
Data File : VX010545.D
Acq On : 28 Jun 2019 19:18
Operator : JC/SP
Sample : K3572-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ERM-7

Quant Time: Jun 29 06:27:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

