

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062819\
 Data File : VX010565.D
 Acq On : 29 Jun 2019 03:51
 Operator : JC/SP
 Sample : K3555-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SAMPLE-18

Quant Time: Jun 29 07:01:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	253516	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	398719	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	363054	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	168590	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122357	51.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.96%	
35) Dibromofluoromethane	5.50	113	119704	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
50) Toluene-d8	8.71	98	476036	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
62) 4-Bromofluorobenzene	11.13	95	160175	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
Target Compounds						
18) Methyl Acetate	2.78	43	2533	1.067	ug/l #	86
20) Methylene Chloride	2.86	84	17426	7.108	ug/l	97
43) Isopropyl Acetate	6.45	43	9931	1.842	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062819\
Data File : VX010565.D
Acq On : 29 Jun 2019 03:51
Operator : JC/SP
Sample : K3555-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SAMPLE-18

Quant Time: Jun 29 07:01:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

