

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062819\
 Data File : VX010572.D
 Acq On : 29 Jun 2019 06:35
 Operator : JC/SP
 Sample : K3158-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OK-02-062719

Quant Time: Jul 01 03:59:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	246890	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	386130	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	360561	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	167194	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	120763	52.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.34%	
35) Dibromofluoromethane	5.50	113	116328	49.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.48%	
50) Toluene-d8	8.71	98	466061	51.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.12%	
62) 4-Bromofluorobenzene	11.13	95	157952	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
Target Compounds						
16) Acetone	2.45	43	7022	6.053	ug/l	97
18) Methyl Acetate	2.78	43	1963	0.849	ug/l	92
20) Methylene Chloride	2.86	84	355308	148.822	ug/l	97
43) Isopropyl Acetate	6.45	43	13959	2.673	ug/l	97
95) Naphthalene	13.83	128	31188	2.666	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062819\
Data File : VX010572.D
Acq On : 29 Jun 2019 06:35
Operator : JC/SP
Sample : K3158-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OK-02-062719

Quant Time: Jul 01 03:59:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

