

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX062819\
 Data File : VX010552.D
 Acq On : 28 Jun 2019 22:24
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 29 06:37:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	74	0.00
2 T	Dichlorodifluoromethane	0.348	0.293	15.8	65	0.00
3 P	Chloromethane	0.405	0.342	15.6	65	0.00
4 C	Vinyl Chloride	0.443	0.377	14.9#	67	0.00
5 T	Bromomethane	0.216	0.212	1.9	78	0.00
6 T	Chloroethane	0.262	0.254	3.1	75	0.00
7 T	Trichlorofluoromethane	0.703	0.662	5.8	72	0.00
8 T	Diethyl Ether	0.252	0.241	4.4	73	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.426	0.383	10.1	70	0.00
10 T	Methyl Iodide	0.608	0.574	5.6	68	0.00
11 T	Tert butyl alcohol	0.123	0.115	6.5	73	0.00
12 CM	1,1-Dichloroethene	0.433	0.388	10.4#	68	0.00
13 T	Acrolein	0.081	0.093	-14.8	91	0.00
14 T	Allyl chloride	0.615	0.564	8.3	68	0.00
15 T	Acrylonitrile	0.243	0.241	0.8	74	0.00
16 T	Acetone	0.235	0.200	14.9	61	0.00
17 T	Carbon Disulfide	1.163	0.885	23.9#	58	0.00
18 T	Methyl Acetate	0.468	0.469	-0.2	76	0.00
19 T	Methyl tert-butyl Ether	1.363	1.306	4.2	73	0.00
20 T	Methylene Chloride	0.484	0.455	6.0	72	0.00
21 T	trans-1,2-Dichloroethene	0.466	0.419	10.1	69	0.00
22 T	Diisopropyl ether	1.286	1.248	3.0	74	0.00
23 T	Vinyl Acetate	1.111	1.081	2.7	71	0.00
24 P	1,1-Dichloroethane	0.744	0.708	4.8	72	0.00
25 T	2-Butanone	0.345	0.329	4.6	70	0.00
26 T	2,2-Dichloropropane	0.646	0.487	24.6#	58	0.00
27 T	cis-1,2-Dichloroethene	0.539	0.507	5.9	71	0.00
28 T	Bromochloromethane	0.356	0.364	-2.2	79	0.00
29 T	Tetrahydrofuran	0.213	0.211	0.9	74	0.00
30 C	Chloroform	0.800	0.772	3.5#	74	0.00
31 T	Cyclohexane	0.677	0.604	10.8	67	0.00
32 T	1,1,1-Trichloroethane	0.706	0.663	6.1	70	0.00
33 S	1,2-Dichloroethane-d4	0.469	0.480	-2.3	76	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	76	0.00
35 S	Dibromofluoromethane	0.306	0.307	-0.3	75	0.00
36 T	1,1-Dichloropropene	0.397	0.347	12.6	68	0.00
37 T	Ethyl Acetate	0.409	0.404	1.2	74	0.00
38 T	Carbon Tetrachloride	0.430	0.378	12.1	67	0.00
39 T	Methylcyclohexane	0.522	0.446	14.6	66	0.00
40 TM	Benzene	1.238	1.149	7.2	71	0.00
41 T	Methacrylonitrile	0.224	0.217	3.1	74	0.00
42 TM	1,2-Dichloroethane	0.377	0.369	2.1	73	0.00
43 T	Isopropyl Acetate	0.676	0.649	4.0	73	0.00
44 TM	Trichloroethene	0.370	0.338	8.6	70	0.00
45 C	1,2-Dichloropropane	0.313	0.290	7.3#	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX062819\
 Data File : VX010552.D
 Acq On : 28 Jun 2019 22:24
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 29 06:37:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.224	0.210	6.3	72	0.00
47 T	Bromodichloromethane	0.411	0.380	7.5	69	0.00
48 T	Methyl methacrylate	0.314	0.308	1.9	73	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	75	0.00
50 S	Toluene-d8	1.170	1.151	1.6	73	0.00
51 T	4-Methyl-2-Pentanone	0.433	0.433	0.0	75	0.00
52 CM	Toluene	0.815	0.747	8.3#	70	0.00
53 T	t-1,3-Dichloropropene	0.461	0.405	12.1	66	0.00
54 T	cis-1,3-Dichloropropene	0.501	0.454	9.4	68	0.00
55 T	1,1,2-Trichloroethane	0.336	0.325	3.3	73	0.00
56 T	Ethyl methacrylate	0.509	0.486	4.5	71	0.00
57 T	1,3-Dichloropropane	0.522	0.499	4.4	73	0.00
58 T	2-Chloroethyl Vinyl ether	0.242	0.240	0.8	73	0.00
59 T	2-Hexanone	0.340	0.333	2.1	73	0.00
60 T	Dibromochloromethane	0.376	0.344	8.5	68	0.00
61 T	1,2-Dibromoethane	0.363	0.342	5.8	71	0.00
62 S	4-Bromofluorobenzene	0.423	0.416	1.7	73	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	75	0.00
64 T	Tetrachloroethene	0.404	0.377	6.7	73	0.00
65 PM	Chlorobenzene	1.015	0.940	7.4	71	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.345	6.8	69	0.00
67 C	Ethyl Benzene	1.714	1.597	6.8#	71	0.00
68 T	m/p-Xylenes	0.664	0.619	6.8	70	0.00
69 T	o-Xylene	0.653	0.613	6.1	72	0.00
70 T	Styrene	1.095	1.043	4.7	71	0.00
71 P	Bromoform	0.322	0.285	11.5	63	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	75	0.00
73 T	Isopropylbenzene	3.389	3.150	7.1	72	0.00
74 T	N-amyl acetate	1.281	1.251	2.3	71	0.00
75 P	1,1,2,2-Tetrachloroethane	1.122	1.059	5.6	72	0.00
76 T	1,2,3-Trichloropropane	0.911	0.873	4.2	72	0.00
77 T	Bromobenzene	0.946	0.865	8.6	71	0.00
78 T	n-propylbenzene	3.725	3.506	5.9	71	0.00
79 T	2-Chlorotoluene	2.218	2.068	6.8	71	0.00
80 T	1,3,5-Trimethylbenzene	2.794	2.620	6.2	70	0.00
81 T	trans-1,4-Dichloro-2-butene	0.391	0.309	21.0#	59	0.00
82 T	4-Chlorotoluene	2.524	2.366	6.3	70	0.00
83 T	tert-Butylbenzene	2.804	2.594	7.5	70	0.00
84 T	1,2,4-Trimethylbenzene	2.807	2.658	5.3	70	0.00
85 T	sec-Butylbenzene	3.293	3.117	5.3	71	0.00
86 T	p-Isopropyltoluene	3.046	2.832	7.0	70	0.00
87 T	1,3-Dichlorobenzene	1.652	1.513	8.4	71	0.00
88 T	1,4-Dichlorobenzene	1.682	1.511	10.2	70	0.00
89 T	n-Butylbenzene	2.588	2.407	7.0	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX062819\
Data File : VX010552.D
Acq On : 28 Jun 2019 22:24
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 29 06:37:55 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.527	0.451	14.4	62	0.00
91 T	1,2-Dichlorobenzene	1.642	1.505	8.3	71	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.242	0.216	10.7	70	0.00
93 T	1,2,4-Trichlorobenzene	1.139	1.041	8.6	69	0.00
94 T	Hexachlorobutadiene	0.556	0.492	11.5	67	0.00
95 T	Naphthalene	3.499	3.323	5.0	70	0.00
96 T	1,2,3-Trichlorobenzene	1.132	1.058	6.5	70	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6