

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062819\
 Data File : VX010526.D
 Acq On : 28 Jun 2019 10:51
 Operator : JC/SP
 Sample : VX0628WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0628WBL01

Quant Time: Jun 29 04:52:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	276082	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	413863	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	363369	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	161626	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	137858	53.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.52%	
35) Dibromofluoromethane	5.49	113	137635	54.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.70%	
50) Toluene-d8	8.71	98	514924	53.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.30%	
62) 4-Bromofluorobenzene	11.13	95	164865	47.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.28%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062819\
Data File : VX010526.D
Acq On : 28 Jun 2019 10:51
Operator : JC/SP
Sample : VX0628WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0628WBL01

Quant Time: Jun 29 04:52:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX010526.D

Acq: 28 Jun 2019 10:51

Instrument :

MSVOA_X

ClientSampled :


VX0628WBL01

Tot Ion:168 Resp: 276082

Ion Ratio Lower Upper

168 100

99 45.9 35.5 53.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

100000

80000

60000



40000

20000

0

Time-->

5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 53.257 ug/l

RT: 6.06 min Scan# 805

Delta R.T. 0.00 min

Lab File: VX010526.D


Acq: 28 Jun 2019 10:51

Tgt Ion: 65 Resp: 137858

Ion Ratio Lower Upper

65 100

67 54.0 0.0 110.2



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06

60000

50000

40000

30000


20000


10000

0

Time-->

5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 935

Delta R.T. -0.01 min

Lab File: VX010526.D

Acq: 28 Jun 2019 10:51

Instrument :

MSVOA_X

ClientSampleId :

VX0628WBL01


Tot Ion:114 Resp: 413863

Ion Ratio Lower Upper

114 100

63 17.2 0.0 33.8




88 14.4 0.0 27.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 54.355 ug/l

RT: 5.49 min Scan# 712

Delta R.T. -0.01 min

Lab File: VX010526.D

Acq: 28 Jun 2019 10:51


Tgt Ion:113 Resp: 137635

Ion Ratio Lower Upper

113 100

111 101.2 81.2 121.8



192 23.4 18.6 27.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 53.153 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX010526.D

Acq: 28 Jun 2019 10:51

Instrument :

MSVOA_X

ClientSampleId :


VX0628WBL01

Tot Ion: 98 Resp: 514924

Ion Ratio Lower Upper

98 100

100 64.7 51.6 77.4



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V

8.71

300000

200000

100000



0

Time-->

8.60

8.70

8.80



#62

4-Bromofluorobenzene

Concen: 47.138 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX010526.D

Acq: 28 Jun 2019 10:51


Tgt Ion: 95 Resp: 164865

Ion Ratio Lower Upper

95 100

174 97.8 0.0 187.6

176 95.8 0.0 184.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

11.13

150000

100000


50000


0

Time-->

11.10

11.20









#63
 Chlorobenzene-d5
 Concen: 50.000 ug/l
 RT: 10.11 min Scan# 1469
 Delta R.T. 0.00 min
 Lab File: VX010526.D
 Acq: 28 Jun 2019 10:51

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0628WBL01

Tot Ion:117 Resp: 363369
 Ion Ratio Lower Upper
 117 100
 82 49.9 41.2 61.8
 119 31.4 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
 Ion 82.00 (81.70 to 82.70): VX0
 Ion 118.90 (118.60 to 119.60): V



#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 12.07 min Scan# 1791
 Delta R.T. -0.01 min
 Lab File: VX010526.D
 Acq: 28 Jun 2019 10:51

Tgt Ion:152 Resp: 161626
 Ion Ratio Lower Upper
 152 100
 115 54.0 38.6 115.7
 150 157.0 0.0 347.4



Abundance Ion 151.90 (151.60 to 152.60): V
 Ion 114.90 (114.60 to 115.60): V
 Ion 149.90 (149.60 to 150.60): V

