

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
 Data File : VX022967.D
 Acq On : 28 Jun 2021 09:59
 Operator : JC/MD
 Sample : VX0628MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0628MBL01

Quant Time: Jun 28 16:19:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	55400	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	111778	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41258	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49068	48.805	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.600%
35) Dibromofluoromethane	5.391	113	35101	49.451	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.900%
50) Toluene-d8	8.653	98	140000	50.325	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.660%
62) 4-Bromofluorobenzene	11.085	95	46983	45.253	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
Data File : VX022967.D
Acq On : 28 Jun 2021 09:59
Operator : JC/MD
Sample : VX0628MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0628MBL01

Quant Time: Jun 28 16:19:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

