

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
 Data File : VX022971.D
 Acq On : 28 Jun 2021 11:54
 Operator : JC/MD
 Sample : M2838-19DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D2-20210622DL

Quant Time: Jun 29 03:04:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	54390	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	109397	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	95164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	33212	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52975	53.669	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.340%
35) Dibromofluoromethane	5.391	113	37191	53.536	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.080%
50) Toluene-d8	8.653	98	146263	53.721	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.440%
62) 4-Bromofluorobenzene	11.085	95	46888	46.144	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.280%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	252	0.357	ug/l	# 84
27) cis-1,2-Dichloroethene	4.501	96	7807	8.986	ug/l	96
30) Chloroform	5.099	83	1043	0.696	ug/l	85
44) Trichloroethene	7.129	130	42421	53.387	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
Data File : VX022971.D
Acq On : 28 Jun 2021 11:54
Operator : JC/MD
Sample : M2838-19DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D2-20210622DL

Quant Time: Jun 29 03:04:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

