

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
 Data File : VX022972.D
 Acq On : 28 Jun 2021 12:17
 Operator : JC/MD
 Sample : M2838-13DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D2-20210622DL

Quant Time: Jun 29 03:04:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	109477	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	94139	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	34172	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53887	55.954	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.900%
35) Dibromofluoromethane	5.403	113	37199	53.508	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.020%
50) Toluene-d8	8.653	98	146459	53.754	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.500%
62) 4-Bromofluorobenzene	11.085	95	47699	46.908	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.820%
Target Compounds						Qvalue
44) Trichloroethene	7.135	130	45536	57.266	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
Data File : VX022972.D
Acq On : 28 Jun 2021 12:17
Operator : JC/MD
Sample : M2838-13DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D2-20210622DL

Quant Time: Jun 29 03:04:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

