

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
 Data File : VX022974.D
 Acq On : 28 Jun 2021 13:04
 Operator : JC/MD
 Sample : M2839-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB04-20210622

Quant Time: Jun 29 03:05:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53242	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	111637	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	100090	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53565	55.437	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.880%
35) Dibromofluoromethane	5.397	113	37101	52.334	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.660%
50) Toluene-d8	8.653	98	148364	53.399	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.800%
62) 4-Bromofluorobenzene	11.085	95	49837	48.062	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.120%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.971	59	5957	29.582	ug/l	96
16) Acetone	2.386	43	11041	25.775	ug/l	90
25) 2-Butanone	4.580	43	5895	8.850	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
Data File : VX022974.D
Acq On : 28 Jun 2021 13:04
Operator : JC/MD
Sample : M2839-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB04-20210622

Quant Time: Jun 29 03:05:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

