

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
 Data File : VX022977.D
 Acq On : 28 Jun 2021 14:15
 Operator : JC/MD
 Sample : M2839-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW4-20210623

Quant Time: Jun 29 03:05:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	51890	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	107222	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	96351	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	34944	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53202	56.496	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.000%
35) Dibromofluoromethane	5.397	113	35929	52.768	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.540%
50) Toluene-d8	8.653	98	144491	54.147	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.300%
62) 4-Bromofluorobenzene	11.085	95	49357	49.559	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.120%
Target Compounds						
44) Trichloroethene	7.135	130	12009	15.420	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
Data File : VX022977.D
Acq On : 28 Jun 2021 14:15
Operator : JC/MD
Sample : M2839-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW4-20210623

Quant Time: Jun 29 03:05:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

