

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
 Data File : VX022979.D
 Acq On : 28 Jun 2021 15:02
 Operator : JC/MD
 Sample : M2839-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB04-20210622

Quant Time: Jun 29 03:06:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	52873	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	111957	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100382	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37921	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54192	56.477	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.960%
35) Dibromofluoromethane	5.397	113	36519	51.366	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.740%
50) Toluene-d8	8.653	98	149143	53.526	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.060%
62) 4-Bromofluorobenzene	11.085	95	50683	48.739	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.480%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.977	59	5291	26.458	ug/l	# 94
16) Acetone	2.386	43	10465	24.601	ug/l	89
25) 2-Butanone	4.580	43	5602	8.469	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
Data File : VX022979.D
Acq On : 28 Jun 2021 15:02
Operator : JC/MD
Sample : M2839-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB04-20210622

Quant Time: Jun 29 03:06:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

