

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
 Data File : VX022981.D
 Acq On : 28 Jun 2021 16:13
 Operator : JC/MD
 Sample : M2824-03RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ABN-1RE

Quant Time: Jun 28 16:35:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	50203	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	102768	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	90753	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	32570	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	46123	50.625	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.240%
35) Dibromofluoromethane	5.397	113	14609	22.386	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	44.780%#
50) Toluene-d8	8.653	98	128247	50.142	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.280%
62) 4-Bromofluorobenzene	11.085	95	42943	44.988	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.980%
Target Compounds						
16) Acetone	2.386	43	20781	51.450	ug/l	95
20) Methylene Chloride	2.794	84	5393	7.054	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
Data File : VX022981.D
Acq On : 28 Jun 2021 16:13
Operator : JC/MD
Sample : M2824-03RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ABN-1RE

Quant Time: Jun 28 16:35:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

