

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
 Data File : VX022986.D
 Acq On : 28 Jun 2021 18:10
 Operator : JC/MD
 Sample : M2875-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Jun 29 03:07:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	57689	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	118968	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	103071	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37546	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52884	50.513	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.020%
35) Dibromofluoromethane	5.397	113	37144	49.166	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.340%
50) Toluene-d8	8.653	98	143867	48.590	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.180%
62) 4-Bromofluorobenzene	11.085	95	49672	44.951	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.900%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	5166	11.130	ug/l	# 88
19) Methyl tert-butyl Ether	3.123	73	1802	0.671	ug/l	# 58
30) Chloroform	5.105	83	7959	5.008	ug/l	94
64) Tetrachloroethene	9.281	164	481	0.664	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
Data File : VX022986.D
Acq On : 28 Jun 2021 18:10
Operator : JC/MD
Sample : M2875-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Jun 29 03:07:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

