

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
 Data File : VX022966.D
 Acq On : 28 Jun 2021 09:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 28 16:18:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.705	0.707	-0.3	99	0.00
3 P	Chloromethane	0.775	0.692	10.7	96	0.00
4 C	Vinyl Chloride	0.715	0.687	3.9#	98	0.00
5 T	Bromomethane	0.394	0.380	3.6	108	0.00
6 T	Chloroethane	0.442	0.404	8.6	94	0.00
7 T	Trichlorofluoromethane	1.038	1.075	-3.6	107	0.00
8 T	Diethyl Ether	0.401	0.386	3.7	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.650	0.683	-5.1	108	0.00
10 T	Methyl Iodide	0.722	0.510	29.4#	72	0.00
11 T	Tert butyl alcohol	0.189	0.170	10.1	97	0.00
12 CM	1,1-Dichloroethene	0.649	0.632	2.6#	102	0.00
13 T	Acrolein	0.118	0.110	6.8	102	0.00
14 T	Allyl chloride	1.232	1.243	-0.9	105	0.00
15 T	Acrylonitrile	0.432	0.438	-1.4	103	0.00
16 T	Acetone	0.402	0.420	-4.5	116	0.00
17 T	Carbon Disulfide	1.635	1.664	-1.8	110	0.00
18 T	Methyl Acetate	0.990	0.994	-0.4	108	0.00
19 T	Methyl tert-butyl Ether	2.326	2.364	-1.6	104	0.00
20 T	Methylene Chloride	0.848	0.756	10.8	102	0.00
21 T	trans-1,2-Dichloroethene	0.720	0.708	1.7	105	0.00
22 T	Diisopropyl ether	2.563	2.603	-1.6	102	0.00
23 T	Vinyl Acetate	2.150	2.274	-5.8	103	0.00
24 P	1,1-Dichloroethane	1.369	1.382	-0.9	104	0.00
25 T	2-Butanone	0.626	0.647	-3.4	106	0.00
26 T	2,2-Dichloropropane	1.039	1.134	-9.1	109	0.00
27 T	cis-1,2-Dichloroethene	0.799	0.804	-0.6	103	0.00
28 T	Bromochloromethane	0.645	0.652	-1.1	106	0.00
29 T	Tetrahydrofuran	0.402	0.401	0.2	100	0.00
30 C	Chloroform	1.378	1.359	1.4#	103	0.00
31 T	Cyclohexane	1.281	1.269	0.9	103	0.00
32 T	1,1,1-Trichloroethane	1.107	1.140	-3.0	104	0.00
33 S	1,2-Dichloroethane-d4	0.907	0.893	1.5	110	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.318	0.328	-3.1	108	0.00
36 T	1,1-Dichloropropene	0.531	0.541	-1.9	106	0.00
37 T	Ethyl Acetate	0.615	0.618	-0.5	102	0.00
38 T	Carbon Tetrachloride	0.445	0.487	-9.4	105	0.00
39 T	Methylcyclohexane	0.620	0.669	-7.9	107	0.00
40 TM	Benzene	1.531	1.584	-3.5	104	0.00
41 T	Methacrylonitrile	0.339	0.346	-2.1	106	0.00
42 TM	1,2-Dichloroethane	0.563	0.587	-4.3	102	0.00
43 T	Isopropyl Acetate	0.989	1.029	-4.0	103	0.00
44 TM	Trichloroethene	0.363	0.369	-1.7	103	0.00
45 C	1,2-Dichloropropane	0.395	0.413	-4.6#	104	0.00
46 T	Dibromomethane	0.262	0.268	-2.3	105	0.00
47 T	Bromodichloromethane	0.466	0.513	-10.1	106	0.00
48 T	Methyl methacrylate	0.458	0.492	-7.4	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
 Data File : VX022966.D
 Acq On : 28 Jun 2021 09:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 28 16:18:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.009	10.0	91	0.00
50 S	Toluene-d8	1.244	1.220	1.9	102	0.00
51 T	4-Methyl-2-Pentanone	0.616	0.635	-3.1	99	0.00
52 CM	Toluene	0.926	0.944	-1.9#	100	0.00
53 T	t-1,3-Dichloropropene	0.528	0.599	-13.4	105	0.00
54 T	cis-1,3-Dichloropropene	0.594	0.658	-10.8	104	0.00
55 T	1,1,2-Trichloroethane	0.371	0.378	-1.9	100	0.00
56 T	Ethyl methacrylate	0.607	0.635	-4.6	98	0.00
57 T	1,3-Dichloropropane	0.675	0.681	-0.9	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.309	0.336	-8.7	101	0.00
59 T	2-Hexanone	0.467	0.512	-9.6	105	0.00
60 T	Dibromochloromethane	0.311	0.344	-10.6	103	0.00
61 T	1,2-Dibromoethane	0.377	0.374	0.8	98	0.00
62 S	4-Bromofluorobenzene	0.464	0.465	-0.2	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.351	0.371	-5.7	102	0.00
65 PM	Chlorobenzene	1.053	1.097	-4.2	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.340	0.368	-8.2	98	0.00
67 C	Ethyl Benzene	1.939	2.084	-7.5#	102	0.00
68 T	m/p-Xylenes	0.720	0.763	-6.0	102	0.00
69 T	o-Xylene	0.701	0.738	-5.3	100	0.00
70 T	Styrene	1.163	1.247	-7.2	101	0.00
71 P	Bromoform	0.194	0.230	-18.6	106	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	4.231	4.324	-2.2	102	0.00
74 T	N-amyl acetate	2.118	2.251	-6.3	102	0.00
75 P	1,1,2,2-Tetrachloroethane	1.453	1.419	2.3	97	0.00
76 T	1,2,3-Trichloropropane	1.321	1.389	-5.1	105	0.00
77 T	Bromobenzene	0.880	0.874	0.7	102	0.00
78 T	n-propylbenzene	5.119	5.501	-7.5	106	0.00
79 T	2-Chlorotoluene	3.091	3.184	-3.0	103	0.00
80 T	1,3,5-Trimethylbenzene	3.474	3.733	-7.5	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.427	0.448	-4.9	104	0.00
82 T	4-Chlorotoluene	3.572	3.744	-4.8	108	0.00
83 T	tert-Butylbenzene	3.281	3.296	-0.5	101	0.00
84 T	1,2,4-Trimethylbenzene	3.496	3.663	-4.8	101	0.00
85 T	sec-Butylbenzene	4.424	4.698	-6.2	104	0.00
86 T	p-Isopropyltoluene	3.516	3.808	-8.3	106	0.00
87 T	1,3-Dichlorobenzene	1.697	1.757	-3.5	105	0.00
88 T	1,4-Dichlorobenzene	1.775	1.775	0.0	106	0.00
89 T	n-Butylbenzene	3.343	3.848	-15.1	111	0.00
90 T	Hexachloroethane	0.549	0.624	-13.7	109	0.00
91 T	1,2-Dichlorobenzene	1.657	1.707	-3.0	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.285	0.317	-11.2	107	0.00
93 T	1,2,4-Trichlorobenzene	0.919	1.030	-12.1	112	0.00
94 T	Hexachlorobutadiene	0.360	0.406	-12.8	122	0.00
95 T	Naphthalene	3.590	3.998	-11.4	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
Data File : VX022966.D
Acq On : 28 Jun 2021 09:27
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 28 16:18:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T 1,2,3-Trichlorobenzene	0.925	0.993	-7.4	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6