

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
 Data File : VX022991.D
 Acq On : 28 Jun 2021 20:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 29 03:09:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	96	0.01
2 T	Dichlorodifluoromethane	50.000	49.114	1.8	89	0.00
3 P	Chloromethane	50.000	48.850	2.3	97	0.00
4 C	Vinyl Chloride	50.000	49.200	1.6#	93	0.00
5 T	Bromomethane	50.000	44.841	10.3	92	0.00
6 T	Chloroethane	50.000	49.431	1.1	93	0.00
7 T	Trichlorofluoromethane	50.000	51.138	-2.3	97	0.00
8 T	Diethyl Ether	50.000	49.992	0.0	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.217	-0.4	95	0.00
10 T	Methyl Iodide	50.000	46.409	7.2	94	0.00
11 T	Tert butyl alcohol	250.000	239.097	4.4	95	0.00
12 CM	1,1-Dichloroethene	50.000	50.152	-0.3#	96	0.00
13 T	Acrolein	250.000	259.375	-3.8	105	0.00
14 T	Allyl chloride	50.000	51.063	-2.1	98	0.00
15 T	Acrylonitrile	250.000	271.949	-8.8	101	0.00
16 T	Acetone	250.000	244.241	2.3	99	0.00
17 T	Carbon Disulfide	50.000	42.983	14.0	91	0.00
18 T	Methyl Acetate	50.000	53.288	-6.6	105	0.00
19 T	Methyl tert-butyl Ether	50.000	52.881	-5.8	100	0.00
20 T	Methylene Chloride	50.000	49.351	1.3	97	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.462	1.1	97	0.00
22 T	Diisopropyl ether	50.000	54.319	-8.6	101	0.00
23 T	Vinyl Acetate	250.000	278.735	-11.5	100	0.00
24 P	1,1-Dichloroethane	50.000	52.164	-4.3	99	0.00
25 T	2-Butanone	250.000	265.515	-6.2	100	0.00
26 T	2,2-Dichloropropane	50.000	45.907	8.2	85	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.808	-3.6	98	0.00
28 T	Bromochloromethane	50.000	52.469	-4.9	102	0.00
29 T	Tetrahydrofuran	250.000	273.689	-9.5	101	0.00
30 C	Chloroform	50.000	51.943	-3.9#	99	0.00
31 T	Cyclohexane	50.000	50.099	-0.2	96	0.00
32 T	1,1,1-Trichloroethane	50.000	53.579	-7.2	100	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.173	1.7	101	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	99	0.00
35 S	Dibromofluoromethane	50.000	49.406	1.2	100	0.00
36 T	1,1-Dichloropropene	50.000	48.113	3.8	97	0.00
37 T	Ethyl Acetate	50.000	52.886	-5.8	104	0.00
38 T	Carbon Tetrachloride	50.000	51.308	-2.6	95	0.00
39 T	Methylcyclohexane	50.000	49.269	1.5	95	0.00
40 TM	Benzene	50.000	50.711	-1.4	99	0.00
41 T	Methacrylonitrile	50.000	54.136	-8.3	108	0.00
42 TM	1,2-Dichloroethane	50.000	53.296	-6.6	101	0.00
43 T	Isopropyl Acetate	50.000	53.159	-6.3	102	0.00
44 TM	Trichloroethene	50.000	49.841	0.3	98	0.00
45 C	1,2-Dichloropropane	50.000	52.062	-4.1#	100	0.00
46 T	Dibromomethane	50.000	51.556	-3.1	102	0.00
47 T	Bromodichloromethane	50.000	47.351	5.3	100	0.00
48 T	Methyl methacrylate	50.000	53.289	-6.6	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
 Data File : VX022991.D
 Acq On : 28 Jun 2021 20:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 29 03:09:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	898.181	10.2	85	0.00
50 S	Toluene-d8	50.000	49.111	1.8	99	0.00
51 T	4-Methyl-2-Pentanone	250.000	270.739	-8.3	100	0.00
52 CM	Toluene	50.000	50.656	-1.3#	96	0.00
53 T	t-1,3-Dichloropropene	50.000	45.538	8.9	97	0.00
54 T	cis-1,3-Dichloropropene	50.000	46.129	7.7	95	0.00
55 T	1,1,2-Trichloroethane	50.000	49.749	0.5	95	0.00
56 T	Ethyl methacrylate	50.000	47.674	4.7	97	0.00
57 T	1,3-Dichloropropane	50.000	51.792	-3.6	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	249.200	0.3	99	0.00
59 T	2-Hexanone	250.000	270.566	-8.2	100	0.00
60 T	Dibromochloromethane	50.000	45.356	9.3	95	0.00
61 T	1,2-Dibromoethane	50.000	51.196	-2.4	98	0.00
62 S	4-Bromofluorobenzene	50.000	51.632	-3.3	106	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	97	0.00
64 T	Tetrachloroethene	50.000	50.576	-1.2	95	0.00
65 PM	Chlorobenzene	50.000	50.094	-0.2	97	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.950	-9.9	98	0.00
67 C	Ethyl Benzene	50.000	52.469	-4.9#	97	0.00
68 T	m/p-Xylenes	100.000	102.565	-2.6	97	0.00
69 T	o-Xylene	50.000	51.390	-2.8	96	0.00
70 T	Styrene	50.000	52.052	-4.1	96	0.00
71 P	Bromoform	50.000	44.203	11.6	97	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	97	0.00
73 T	Isopropylbenzene	50.000	52.623	-5.2	97	0.00
74 T	N-ethyl acetate	50.000	50.614	-1.2	101	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	52.947	-5.9	98	0.00
76 T	1,2,3-Trichloropropane	50.000	53.528	-7.1	99	0.00
77 T	Bromobenzene	50.000	51.109	-2.2	97	0.00
78 T	n-propylbenzene	50.000	53.647	-7.3	98	0.00
79 T	2-Chlorotoluene	50.000	53.284	-6.6	99	0.00
80 T	1,3,5-Trimethylbenzene	50.000	55.033	-10.1	100	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.150	9.7	94	0.00
82 T	4-Chlorotoluene	50.000	53.194	-6.4	102	0.00
83 T	tert-Butylbenzene	50.000	52.068	-4.1	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	55.142	-10.3	99	0.00
85 T	sec-Butylbenzene	50.000	53.642	-7.3	98	0.00
86 T	p-Isopropyltoluene	50.000	52.756	-5.5	96	0.00
87 T	1,3-Dichlorobenzene	50.000	50.938	-1.9	96	0.00
88 T	1,4-Dichlorobenzene	50.000	49.005	2.0	96	0.00
89 T	n-Butylbenzene	50.000	56.137	-12.3	100	0.00
90 T	Hexachloroethane	50.000	45.823	8.4	96	0.00
91 T	1,2-Dichlorobenzene	50.000	52.290	-4.6	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.863	0.3	107	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.241	-4.5	97	0.00
94 T	Hexachlorobutadiene	50.000	48.887	2.2	98	0.00
95 T	Naphthalene	50.000	49.869	0.3	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062821\
Data File : VX022991.D
Acq On : 28 Jun 2021 20:08
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 29 03:09:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	54.066	-8.1	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6