

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062822\
 Data File : VX029839.D
 Acq On : 28 Jun 2022 19:26
 Operator : JC/MD
 Sample : N3494-04
 Misc : 6.77g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022
 Supervised By :Semsettin Yesilyurt 06/30/2022

Quant Time: Jun 29 06:07:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	243784	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	411430	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	344128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153276	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	131749	42.694	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	85.380%		
35) Dibromofluoromethane	5.391	113	114132	44.388	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	88.780%		
50) Toluene-d8	8.653	98	465484	48.257	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.520%		
62) 4-Bromofluorobenzene	11.085	95	150902	44.257	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	88.520%		
Target Compounds						Qvalue
18) Methyl Acetate	2.727	43	12483	2.682	ug/l	97
31) Cyclohexane	5.464	56	5053	1.174	ug/l #	77
39) Methylcyclohexane	7.373	83	11472	2.774	ug/l #	91
40) Benzene	6.031	78	68104	6.159	ug/l	98
52) Toluene	8.720	92	24035	3.602	ug/l	96
67) Ethyl Benzene	10.195	91	34082	2.837	ug/l	98
68) m/p-Xylenes	10.305	106	34704	7.531	ug/l	95
73) Isopropylbenzene	10.963	105	12699	1.145	ug/l	100
78) n-propylbenzene	11.311	91	71466	5.463	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	27994	3.035	ug/l	99
89) n-Butylbenzene	12.335	91	41905	4.857	ug/l	76
95) Naphthalene	13.780	128	43061	3.478	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062822\
Data File : VX029839.D
Acq On : 28 Jun 2022 19:26
Operator : JC/MD
Sample : N3494-04
Misc : 6.77g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jun 29 06:07:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022
Supervised By :Semsettin Yesilyurt 06/30/2022

