

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX062822\  
 Data File : VX029845.D  
 Acq On : 28 Jun 2022 21:46  
 Operator : JC/MD  
 Sample : N3494-34  
 Misc : 5.67g/5.0mL/100uL/5.0mL/MSVOA\_X/MEOH  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SP-EXC-1-1-GR

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022  
 Supervised By :Semsettin Yesilyurt 06/30/2022

Quant Time: Jun 29 06:09:40 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X062822W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jun 29 05:54:51 2022  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response   | Conc    | Units | Dev(Min) |
|-----------------------------|----------------|------|------------|---------|-------|----------|
| -----                       |                |      |            |         |       |          |
| Internal Standards          |                |      |            |         |       |          |
| 1) Pentafluorobenzene       | 5.550          | 168  | 263143     | 50.000  | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763          | 114  | 440532     | 50.000  | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055         | 117  | 372354     | 50.000  | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024         | 152  | 166062     | 50.000  | ug/l  | 0.00     |
|                             |                |      |            |         |       |          |
| System Monitoring Compounds |                |      |            |         |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958          | 65   | 138159     | 41.477  | ug/l  | 0.00     |
| Spiked Amount 50.000        | Range 74 - 125 |      | Recovery = | 82.960% |       |          |
| 35) Dibromofluoromethane    | 5.391          | 113  | 119350     | 43.351  | ug/l  | 0.00     |
| Spiked Amount 50.000        | Range 75 - 124 |      | Recovery = | 86.700% |       |          |
| 50) Toluene-d8              | 8.647          | 98   | 500690     | 48.478  | ug/l  | 0.00     |
| Spiked Amount 50.000        | Range 86 - 113 |      | Recovery = | 96.960% |       |          |
| 62) 4-Bromofluorobenzene    | 11.085         | 95   | 162038     | 44.384  | ug/l  | 0.00     |
| Spiked Amount 50.000        | Range 83 - 123 |      | Recovery = | 88.760% |       |          |
|                             |                |      |            |         |       |          |
| Target Compounds            |                |      |            |         |       |          |
| 18) Methyl Acetate          | 2.727          | 43   | 5234       | 1.042   | ug/l  | 98       |
| 65) Chlorobenzene           | 10.080         | 112  | 51430      | 6.881   | ug/l  | 99       |
| 88) 1,4-Dichlorobenzene     | 12.049         | 146  | 13278      | 2.345   | ug/l  | 79       |
| 91) 1,2-Dichlorobenzene     | 12.341         | 146  | 90120      | 16.531  | ug/l  | 99       |
| 95) Naphthalene             | 13.780         | 128  | 24593m     | 1.834   | ug/l  |          |
| -----                       |                |      |            |         |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX062822\  
Data File : VX029845.D  
Acq On : 28 Jun 2022 21:46  
Operator : JC/MD  
Sample : N3494-34  
Misc : 5.67g/5.0mL/100uL/5.0mL/MSVOA\_X/MEOH  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
SP-EXC-1-1-GR

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 06/30/2022  
Supervised By :Semsettin Yesilyurt 06/30/2022

Quant Time: Jun 29 06:09:40 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X062822W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jun 29 05:54:51 2022  
Response via : Initial Calibration

