

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062823\
 Data File : VX036357.D
 Acq On : 28 Jun 2023 10:26
 Operator : JC/MD
 Sample : VX0628MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0628MBL01

Quant Time: Jun 29 01:25:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	246175	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	428683	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	389714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	180978	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	193110	54.725	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.440%
35) Dibromofluoromethane	5.379	113	148122	52.403	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.800%
50) Toluene-d8	8.647	98	536283	50.797	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.600%
62) 4-Bromofluorobenzene	11.079	95	200041	53.794	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062823\
Data File : VX036357.D
Acq On : 28 Jun 2023 10:26
Operator : JC/MD
Sample : VX0628MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0628MBL01

Quant Time: Jun 29 01:25:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

