

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062823\
 Data File : VX036363.D
 Acq On : 28 Jun 2023 13:08
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 29 01:28:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	189386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	336914	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	303735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143838	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	156258	57.560	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.120%
35) Dibromofluoromethane	5.385	113	118773	53.465	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.920%
50) Toluene-d8	8.646	98	406473	48.988	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.980%
62) 4-Bromofluorobenzene	11.079	95	149695	51.220	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062823\
Data File : VX036363.D
Acq On : 28 Jun 2023 13:08
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 29 01:28:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

