

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062918\
 Data File : VX003027.D
 Acq On : 29 Jun 2018 14:31
 Operator : JC/MD
 Sample : J3743-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GMW-5

Quant Time: Jun 29 15:29:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	264760	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	371349	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	341944	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	193009	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	176264	48.17	ug/l	0.00
Spiked Amount			Recovery	=	96.34%	
35) Dibromofluoromethane	5.51	113	142185	48.35	ug/l	0.00
Spiked Amount			Recovery	=	96.70%	
50) Toluene-d8	8.72	98	516145	48.06	ug/l	0.00
Spiked Amount			Recovery	=	96.12%	
62) 4-Bromofluorobenzene	11.14	95	182489	44.06	ug/l	0.00
Spiked Amount			Recovery	=	88.12%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	35324	23.31	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062918\
Data File : VX003027.D
Acq On : 29 Jun 2018 14:31
Operator : JC/MD
Sample : J3743-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GMW-5

Quant Time: Jun 29 15:29:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

