

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062918\
 Data File : VX003034.D
 Acq On : 29 Jun 2018 17:38
 Operator : JC/MD
 Sample : J3695-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5

Quant Time: Jun 30 09:12:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	266514	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	369817	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	348482	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	196571	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	177566	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
35) Dibromofluoromethane	5.51	113	143076	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
50) Toluene-d8	8.72	98	519740	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
62) 4-Bromofluorobenzene	11.14	95	185332	44.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.88%	
Target Compounds						
16) Acetone	2.45	43	31997	20.97	ug/l	98
25) 2-Butanone	4.72	43	6150	2.72	ug/l	96
29) Tetrahydrofuran	5.18	42	7792	5.37	ug/l	98
95) Naphthalene	13.84	128	9253	0.69	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062918\
Data File : VX003034.D
Acq On : 29 Jun 2018 17:38
Operator : JC/MD
Sample : J3695-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5

Quant Time: Jun 30 09:12:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

