

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062918\
 Data File : VX003035.D
 Acq On : 29 Jun 2018 18:04
 Operator : JC/MD
 Sample : J3695-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LIC-FIELDBLANK-1-06-20-18

Quant Time: Jun 30 09:13:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	263909	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	368235	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	339562	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	190219	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	173201	47.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.96%	
35) Dibromofluoromethane	5.51	113	139948	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
50) Toluene-d8	8.72	98	516023	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
62) 4-Bromofluorobenzene	11.14	95	179727	43.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.52%	
Target Compounds						
16) Acetone	2.45	43	20712	13.71	ug/l	98
20) Methylene Chloride	2.86	84	4909	1.67	ug/l	92
65) Chlorobenzene	10.14	112	2242	0.26	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062918\
Data File : VX003035.D
Acq On : 29 Jun 2018 18:04
Operator : JC/MD
Sample : J3695-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LIC-FIELDBLANK-1-06-20-18

Quant Time: Jun 30 09:13:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

